

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110422\
 Data File : PR057900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Nov 2022 18:42
 Operator : AJ\MA
 Sample : PB148821BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ALCS821

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 07 00:32:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110122CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 01 11:10:38 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.921	47012432	33729737	20.438	19.169
2) SA Decachlor...	9.655	8.171	95412502	127.2E6	45.889	47.995
Target Compounds						
3) L1 AR-1016-1	4.925	4.036	8570546	5170625	115.571	100.311
4) L1 AR-1016-2	4.948	4.053	11457987	8171660	116.520	100.657
5) L1 AR-1016-3	5.012	4.232	7347733	4557665	115.779	100.063
6) L1 AR-1016-4	5.115	4.283	6254084	2802780	117.366	102.944
7) L1 AR-1016-5	5.434	4.499	6121802	4060730	114.450	97.935
31) L7 AR-1260-1	6.655	5.588	13002866	8517988	113.238	101.740
32) L7 AR-1260-2	6.939	5.792	15646485	10632582	110.501	102.282
33) L7 AR-1260-3	7.332	5.950	10170815	10323005	94.981	100.011
34) L7 AR-1260-4	7.575	6.457	12415069	8105789	100.885	88.720
35) L7 AR-1260-5	7.913	6.721	22386979	23378824	90.942	93.326

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110422\
Data File : PR057900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Nov 2022 18:42
Operator : AJ\MA
Sample : PB148821BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ALCS821

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 07 00:32:42 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110122CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 01 11:10:38 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

