

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110524\
 Data File : PR069019.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2024 00:09
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543339

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/06/2024
 Supervised By :Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 03:25:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 29 14:51:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.643	2.973	18037164	117.5E6	18.143	19.586
2) SA Decachlor...	9.529	8.256	12945958	89867767	38.832	36.988
Target Compounds						
26) L6 AR-1254-1	5.770	4.936	13315676	137.7E6	397.595	386.782m
27) L6 AR-1254-2	6.006	5.096	19571325	111.3E6	385.647	359.418m
28) L6 AR-1254-3	6.398	5.520	20401285	161.6E6	394.090	386.643m
29) L6 AR-1254-4	6.710	5.768	12193162	92268522	348.546	402.133m
30) L6 AR-1254-5	7.164	6.215	14410030	127.6E6	377.284	397.768

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110524\
Data File : PR069019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2024 00:09
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543339

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/06/2024
Supervised By :Ankita Jodhani 11/06/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 03:25:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 29 14:51:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

