

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110619\
 Data File : PR042899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Nov 2019 01:53
 Operator : SM\AJ
 Sample : K5687-10
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 26A

Manual Integrations
 APPROVED

Ankita
 11/6/2019 1:49:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 06 06:59:24 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 01 09:55:27 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.710	3.971	93265988	336.8E6	15.823m	16.799
2)	SA Decachlor...	10.621	9.063	45291682	195.6E6	11.318	15.562m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR110619\
Data File : PR042899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Nov 2019 01:53
Operator : SM\AJ
Sample : K5687-10
Misc :
ALS Vial : 60 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampled :
26A


Manual Integrations
APPROVED

Ankita
11/6/2019 1:49:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 06 06:59:24 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110119.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 01 09:55:27 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.710 min
Delta R.T.: 0.000 min
Response: 93265988
Conc: 15.82 ng/ml m

Instrument :
ECD_R
ClientSampleId :
26A


Manual Integrations
APPROVED

Ankita
11/6/2019 1:49:34 PM




#1 Tetrachloro-m-xylene

R.T.: 3.971 min
Delta R.T.: 0.001 min
Response: 336836529
Conc: 16.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.621 min
Delta R.T.: 0.000 min
Response: 45291682
Conc: 11.32 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.063 min
Delta R.T.: -0.007 min
Response: 195556079
Conc: 15.56 ng/ml m