

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111124\
 Data File : PR069160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 06:00
 Operator : AJ\MA
 Sample : P4751-18
 Misc :
 ALS Vial : 78 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BH9F8

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/14/2024
 Supervised By :Ankita Jodhani 11/14/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 13 03:57:38 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 29 14:51:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.640	2.971	19407144	122.0E6	19.522	20.325
2) SA Decachlor...	9.530	8.256	14311254	90515077	42.927	37.254
Target Compounds						
26) L6 AR-1254-1	5.768	4.933	2199964	23020480	65.689	64.645m
27) L6 AR-1254-2	6.004	5.094	4549366	21550967	89.644	69.563m
28) L6 AR-1254-3	6.396	5.517	7564424	30289764	146.121	72.469m#
29) L6 AR-1254-4	6.706	5.766	3044931	12327061	87.041m	53.725m#
30) L6 AR-1254-5	7.162	6.214	6104111	120.4E6	159.818m	375.207 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111124\
Data File : PR069160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 06:00
Operator : AJ\MA
Sample : P4751-18
Misc :
ALS Vial : 78 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
BH9F8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/14/2024
Supervised By :Ankita Jodhani 11/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 13 03:57:38 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 29 14:51:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

