

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111224\
 Data File : PR069184.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Nov 2024 12:26
 Operator : AJ\MA
 Sample : P4761-11MSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A0AN0MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/13/2024
 Supervised By :Ankita Jodhani 11/13/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 12 23:57:16 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Oct 29 14:51:31 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.641	2.973	18799320	109.2E6	18.910	18.203
2) SA Decachlor...	9.528	8.254	13356379	92945795	40.063	38.254m
Target Compounds						
3) L1 AR-1016-1	4.861	4.094	14155216	84331533	479.723	455.130
4) L1 AR-1016-2	4.883	4.113	20520208	122.7E6	446.082	458.819
5) L1 AR-1016-3	4.947	4.292	12974826	65316751	440.910	453.901
6) L1 AR-1016-4	5.049	4.343	10601222	48882614	457.276	414.573
7) L1 AR-1016-5	5.364	4.563	9385994	66446691	403.314	412.198
31) L7 AR-1260-1	6.575	5.661	17488203	127.7E6	438.723	456.868
32) L7 AR-1260-2	6.859	5.867	20746882	143.6E6	455.995	429.860
33) L7 AR-1260-3	7.249	6.025	12485628	131.1E6	361.633	457.506 #
34) L7 AR-1260-4	7.491	6.534	15648184	92280511	404.055	394.738
35) L7 AR-1260-5	7.829	6.799	28461866	220.0E6	407.476	407.064

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111224\
Data File : PR069184.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Nov 2024 12:26
Operator : AJ\MA
Sample : P4761-11MSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A0AN0MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/13/2024
Supervised By :Ankita Jodhani 11/13/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 12 23:57:16 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR102924CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Oct 29 14:51:31 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

