

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 05:36:46 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033789.D 1600 =PR033791.D 800 =PR033790.D
 200 =PR033788.D 100 =PR033787.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.492	1.384	1.241	1.486	1.594	1.439	E6	9.27
2)	SA	Decachlorobiphenyl	1.597	1.474	1.468	1.748	1.957	1.649	E6	12.52
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1	4.135	3.430	3.430	4.436	4.775	4.041	E4	14.90
22)	L5	AR-1248-2	5.951	4.748	4.883	6.469	7.181	5.846	E4	17.77
23)	L5	AR-1248-3	6.445	5.380	5.420	6.808	7.486	6.308	E4	14.41
24)	L5	AR-1248-4	7.392	6.265	6.337	7.676	8.313	7.197	E4	12.27
25)	L5	AR-1248-5	6.923	5.847	6.046	7.234	7.743	6.758	E4	11.84
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR033789.D 1600 =PR033791.D 800 =PR033790.D
 200 =PR033788.D 100 =PR033787.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.822	2.697	2.347	2.853	2.929	2.730 E6	8.42	
2) SA	Decachlorobiphenyl	2.698	2.728	2.569	2.889	3.202	2.817 E6	8.64	
3) L1	AR-1016-1						0.000	-1.00	
4) L1	AR-1016-2						0.000	-1.00	
5) L1	AR-1016-3						0.000	-1.00	

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 05:36:46 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033789.D 1600 =PR033791.D 800 =PR033790.D
 200 =PR033788.D 100 =PR033787.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	7.001	6.190	5.911	7.351	7.508	6.792 E4	10.43
22) L5	AR-1248-2	0.924	0.801	0.778	0.967	1.007	0.895 E5	11.32
23) L5	AR-1248-3	0.953	0.828	0.805	0.992	1.030	0.921 E5	10.86
24) L5	AR-1248-4	1.181	1.046	1.010	1.207	1.266	1.142 E5	9.55
25) L5	AR-1248-5	1.190	1.060	1.041	1.201	1.240	1.146 E5	7.82
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range