

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 04:54:19 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033794.D 1600 =PR033796.D 800 =PR033795.D
 200 =PR033793.D 100 =PR033792.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.396	1.511	1.259	1.324	1.454	1.389	E6	7.23
2)	SA	Decachlorobiphenyl	1.597	1.410	1.528	1.811	1.991	1.667	E6	13.94
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	6.441	6.621	6.750	8.148	9.150	7.422	E4	15.90
27)	L6	AR-1254-2	0.949	0.991	1.017	1.249	1.393	1.120	E5	17.17
28)	L6	AR-1254-3	0.957	1.033	1.069	1.287	1.415	1.152	E5	16.59
29)	L6	AR-1254-4	0.722	0.736	0.788	0.978	1.066	0.858	E5	18.06
30)	L6	AR-1254-5	0.881	0.906	0.991	1.218	1.314	1.062	E5	18.24
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR033794.D 1600 =PR033796.D 800 =PR033795.D
 200 =PR033793.D 100 =PR033792.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.669	3.045	2.411	2.504	2.743	2.675	E6	9.17
2) SA	Decachlorobiphenyl	2.691	2.599	2.698	2.966	3.211	2.833	E6	8.88
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 04:54:19 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033794.D 1600 =PR033796.D 800 =PR033795.D
 200 =PR033793.D 100 =PR033792.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	1.550	1.777	1.672	1.892	2.087	1.796 E5	11.48
27) L6	AR-1254-2	1.322	1.492	1.449	1.671	1.858	1.558 E5	13.40
28) L6	AR-1254-3	2.172	2.511	2.485	2.795	3.039	2.600 E5	12.69
29) L6	AR-1254-4	1.374	1.538	1.569	1.808	1.957	1.649 E5	14.04
30) L6	AR-1254-5	1.846	2.004	2.115	2.432	2.610	2.202 E5	14.25
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range