

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 06:54:44 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033799.D 1600 =PR033801.D 800 =PR033800.D
 200 =PR033798.D 100 =PR033797.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.503	1.382	1.310	1.495	1.432	1.425	E6	5.67
2) SA	Decachlorobiphenyl	1.548	1.438	1.408	1.838	2.037	1.654	E6	16.54
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	1.188	1.107	0.993	1.194	1.489	1.194	E5	15.39
37) L8	AR-1262-2	1.996	1.988	1.751	1.992	2.479	2.041	E5	13.04
38) L8	AR-1262-3	1.339	1.281	1.179	1.435	1.717	1.390	E5	14.75
39) L8	AR-1262-4	1.003	0.949	0.880	1.044	1.265	1.028	E5	14.19
40) L8	AR-1262-5	6.808	6.581	6.292	7.683	8.701	7.213	E4	13.59
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR033799.D 1600 =PR033801.D 800 =PR033800.D
 200 =PR033798.D 100 =PR033797.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.909	2.673	2.509	2.900	2.589	2.716	E6	6.68
2) SA	Decachlorobiphenyl	2.576	2.636	2.466	2.903	3.283	2.773	E6	11.81
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR111518CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Nov 16 06:54:44 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR033799.D 1600 =PR033801.D 800 =PR033800.D
 200 =PR033798.D 100 =PR033797.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	2.302	2.303	2.033	2.217	2.724	2.316 E5	10.94
37) L8	AR-1262-2	4.075	4.259	3.709	4.060	4.716	4.164 E5	8.82
38) L8	AR-1262-3	1.525	1.535	1.364	1.584	1.878	1.577 E5	11.89
39) L8	AR-1262-4	2.889	3.014	2.653	2.964	3.454	2.995 E5	9.74
40) L8	AR-1262-5	1.261	1.278	1.190	1.388	1.594	1.342 E5	11.74
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range