

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111519\
 Data File : PR043119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2019 14:51
 Operator : SM\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 02:36:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Nov 16 02:36:29 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.710	3.970	157.6E6	519.1E6	50.000	50.000
2)	SA Decachlor...	10.623	9.065	185.1E6	479.4E6	50.000	50.000
Target Compounds							
36)	L8 AR-1262-1	8.110	7.162	140.3E6	185.3E6	500.000	500.000
37)	L8 AR-1262-2	8.686	7.609	263.6E6	896.7E6	500.000	500.000
38)	L8 AR-1262-3	9.016	7.894	167.8E6	338.2E6	500.000	500.000
39)	L8 AR-1262-4	9.114	7.959	64661152	592.2E6	500.000	500.000
40)	L8 AR-1262-5	9.816	8.473	81516049	255.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111519\
Data File : PR043119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2019 14:51
Operator : SM\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 02:36:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111519.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Nov 16 02:36:29 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

