

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
 Data File : PR058029.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2022 21:33
 Operator : AJ\MA
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 04:42:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 16 04:40:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.917	51878442	73835820	25.637	23.932
2) SA Decachlor...	9.668	8.165	62025614	77320789	26.796	25.221
Target Compounds						
21) L5 AR-1248-1	4.926	4.032	12270390	16018785	269.065	258.342
22) L5 AR-1248-2	5.221	4.279	16024388	21069741	270.746	261.996
23) L5 AR-1248-3	5.435	4.320	19238049	21031475	269.444	260.011
24) L5 AR-1248-4	5.872	4.495	22156859	26087081	266.045	258.995
25) L5 AR-1248-5	5.912	4.906	21596810	26436941	267.689	259.770

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
Data File : PR058029.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2022 21:33
Operator : AJ\MA
Sample : AR1248ICC250
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 04:42:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 16 04:40:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

