

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
 Data File : PR058037.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2022 23:30
 Operator : AJ\MA
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 05:06:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 16 05:04:32 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.917	206.6E6	324.5E6	95.360	99.069
2) SA Decachlor...	9.669	8.165	381.7E6	549.5E6	94.520	99.006
Target Compounds						
41) L9 AR-1268-1	8.231	7.022	312.6E6	482.8E6	955.116	955.756
42) L9 AR-1268-2	8.323	7.091	283.9E6	445.8E6	952.456	985.511
43) L9 AR-1268-3	8.544	7.313	239.0E6	374.7E6	956.231	997.058
44) L9 AR-1268-4	8.956	7.626	109.2E6	156.7E6	935.083	983.989
45) L9 AR-1268-5	9.351	7.921	808.2E6	1265.3E6	978.590	998.018

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
Data File : PR058037.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2022 23:30
Operator : AJ\MA
Sample : AR1268ICC1000
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 05:06:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 16 05:04:32 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

