

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
 Data File : PR058030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Nov 2022 21:47
 Operator : AJ\MA
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 04:43:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 16 04:40:55 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.693	2.917	10064706	12708053	4.974	4.119
2) SA Decachlor...	9.668	8.166	12372260	13619301	5.345	4.442
Target Compounds						
21) L5 AR-1248-1	4.926	4.032	2388540	3243157	52.376	52.304
22) L5 AR-1248-2	5.221	4.279	3142362	4172380	53.093	51.882
23) L5 AR-1248-3	5.435	4.320	3677503	4081781	51.506	50.463
24) L5 AR-1248-4	5.872	4.495	4290838	4879198	51.522	48.441
25) L5 AR-1248-5	5.912	4.906	4147503	4990624	51.408	49.038



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR11522\
Data File : PR058030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Nov 2022 21:47
Operator : AJ\MA
Sample : AR1248ICC050
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 04:43:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 16 04:40:55 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.693 min
Delta R.T.: 0.000 min
Response: 10064706
Conc: 4.97 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 2.917 min
Delta R.T.: 0.000 min
Response: 12708053
Conc: 4.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.668 min
Delta R.T.: 0.002 min
Response: 12372260
Conc: 5.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.166 min
Delta R.T.: 0.000 min
Response: 13619301
Conc: 4.44 ng/ml



#21 AR-1248-1


R.T.: 4.926 min
Delta R.T.: 0.000 min
Response: 2388540
Conc: 52.38 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.032 min
Delta R.T.: 0.000 min
Response: 3243157
Conc: 52.30 ng/ml




#22 AR-1248-2

R.T.: 5.221 min
Delta R.T.: 0.000 min
Response: 3142362
Conc: 53.09 ng/ml



#22 AR-1248-2


R.T.: 4.279 min
Delta R.T.: 0.000 min
Response: 4172380
Conc: 51.88 ng/ml



#23 AR-1248-3


R.T.: 5.435 min
Delta R.T.: 0.000 min
Response: 3677503
Conc: 51.51 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 4.320 min
Delta R.T.: 0.000 min
Response: 4081781
Conc: 50.46 ng/ml




#24 AR-1248-4

R.T.: 5.872 min
Delta R.T.: 0.000 min
Response: 4290838
Conc: 51.52 ng/ml



#24 AR-1248-4


R.T.: 4.495 min
Delta R.T.: 0.000 min
Response: 4879198
Conc: 48.44 ng/ml



#25 AR-1248-5

R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 4147503
Conc: 51.41 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 4990624
Conc: 49.04 ng/ml