

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
 Data File : PR058045.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Nov 2022 01:27
 Operator : AJ\MA
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR111522AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 05:26:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 16 05:22:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.694	2.917	104.9E6	156.1E6	54.272	55.385
2)	SA Decachlor...	9.667	8.163	119.0E6	153.9E6	52.890	53.771
Target Compounds							
21)	L5 AR-1248-1	4.927	4.031	23285499	31917921	511.993	513.643
22)	L5 AR-1248-2	5.222	4.279	30440752	41455764	515.526	515.950
23)	L5 AR-1248-3	5.436	4.320	36541841	41361700	515.198	516.449
24)	L5 AR-1248-4	5.873	4.494	42173023	51589497	508.465	518.904
25)	L5 AR-1248-5	5.912	4.905	41032609	52199076	509.942	516.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
Data File : PR058045.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Nov 2022 01:27
Operator : AJ\MA
Sample : AR1248ICV500
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 05:26:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 16 05:22:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.694 min
Delta R.T.: 0.000 min
Response: 104853482
Conc: 54.27 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1248




#1 Tetrachloro-m-xylene

R.T.: 2.917 min
Delta R.T.: 0.000 min
Response: 156097620
Conc: 55.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.667 min
Delta R.T.: -0.001 min
Response: 118952177
Conc: 52.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.163 min
Delta R.T.: -0.005 min
Response: 153923749
Conc: 53.77 ng/ml



#21 AR-1248-1


R.T.: 4.927 min
Delta R.T.: 0.000 min
Response: 23285499
Conc: 511.99 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1248




#21 AR-1248-1

R.T.: 4.031 min
Delta R.T.: 0.000 min
Response: 31917921
Conc: 513.64 ng/ml




#22 AR-1248-2

R.T.: 5.222 min
Delta R.T.: 0.000 min
Response: 30440752
Conc: 515.53 ng/ml



#22 AR-1248-2


R.T.: 4.279 min
Delta R.T.: 0.000 min
Response: 41455764
Conc: 515.95 ng/ml



#23 AR-1248-3


R.T.: 5.436 min
Delta R.T.: 0.000 min
Response: 36541841
Conc: 515.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1248




#23 AR-1248-3

R.T.: 4.320 min
Delta R.T.: 0.000 min
Response: 41361700
Conc: 516.45 ng/ml




#24 AR-1248-4

R.T.: 5.873 min
Delta R.T.: 0.002 min
Response: 42173023
Conc: 508.46 ng/ml



#24 AR-1248-4


R.T.: 4.494 min
Delta R.T.: 0.000 min
Response: 51589497
Conc: 518.90 ng/ml



#25 AR-1248-5

R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 41032609
Conc: 509.94 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1248



#25 AR-1248-5

R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 52199076
Conc: 516.27 ng/ml