

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
 Data File : PR058046.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Nov 2022 01:42
 Operator : AJ\MA
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 ICVPR111522AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 16 08:50:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 16 05:22:04 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.695	2.917	108.8E6	161.9E6	56.337	57.447
2)	SA Decachlor...	9.667	8.163	120.5E6	156.9E6	53.560	54.814
Target Compounds							
26)	L6 AR-1254-1	5.844	4.863	43673522	81534802	501.725	502.830
27)	L6 AR-1254-2	6.083	5.022	69093482	72562478	501.753	502.118
28)	L6 AR-1254-3	6.478	5.442	75759256	119.0E6	505.686	508.511
29)	L6 AR-1254-4	6.791	5.688	60449945	77268432	504.987	511.630
30)	L6 AR-1254-5	7.250	6.132	68368015	113.7E6	504.290	507.372



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111522\
Data File : PR058046.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Nov 2022 01:42
Operator : AJ\MA
Sample : AR1254ICV500
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 16 08:50:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111522.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 16 05:22:04 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.695 min
Delta R.T.: 0.000 min
Response: 108842125
Conc: 56.34 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1254




#1 Tetrachloro-m-xylene

R.T.: 2.917 min
Delta R.T.: 0.000 min
Response: 161909669
Conc: 57.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.667 min
Delta R.T.: -0.002 min
Response: 120458135
Conc: 53.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.163 min
Delta R.T.: -0.005 min
Response: 156910550
Conc: 54.81 ng/ml



#26 AR-1254-1


R.T.: 5.844 min
Delta R.T.: 0.000 min
Response: 43673522
Conc: 501.72 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1254




#26 AR-1254-1

R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 81534802
Conc: 502.83 ng/ml




#27 AR-1254-2

R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 69093482
Conc: 501.75 ng/ml



#27 AR-1254-2


R.T.: 5.022 min
Delta R.T.: 0.000 min
Response: 72562478
Conc: 502.12 ng/ml



#28 AR-1254-3


R.T.: 6.478 min
Delta R.T.: 0.000 min
Response: 75759256
Conc: 505.69 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1254




#28 AR-1254-3

R.T.: 5.442 min
Delta R.T.: -0.001 min
Response: 119000541
Conc: 508.51 ng/ml




#29 AR-1254-4

R.T.: 6.791 min
Delta R.T.: 0.000 min
Response: 60449945
Conc: 504.99 ng/ml



#29 AR-1254-4


R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 77268432
Conc: 511.63 ng/ml



#30 AR-1254-5

R.T.: 7.250 min
Delta R.T.: -0.001 min
Response: 68368015
Conc: 504.29 ng/ml

Instrument :
ECD_R
ClientSampleId :
ICVPR111522AR1254



#30 AR-1254-5

R.T.: 6.132 min
Delta R.T.: -0.001 min
Response: 113676010
Conc: 507.37 ng/ml