

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111821\
 Data File : PR052651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Nov 2021 12:06
 Operator : AJ\MA
 Sample : M4669-09DL 50X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 A4272DL

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/22/2021
 Supervised By :Sohil Jodhani 11/22/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 01:20:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110921CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 10 04:57:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.673min 32.678 ng/ml

response 117596888

(1) Tetrachloro-m-xylene #2 (SA)

3.852min 0.291 ng/ml

response 856947

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111821\
Data File : PR052651.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Nov 2021 12:06
Operator : AJ\MA
Sample : M4669-09DL 50X
Misc :
ALS Vial : 11 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
A4272DL

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/22/2021
Supervised By :Sohil Jodhani 11/22/2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 19 01:41:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110921CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Nov 10 04:57:17 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

4.673min 32.419 ng/ml m

response 116666221

(1) Tetrachloro-m-xylene #2 (SA)

3.852min 0.291 ng/ml

response 856947

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR111821\
 Data File : PR052651.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Nov 2021 12:06
 Operator : AJ\MA
 Sample : M4669-09DL 50X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A4272DL

Manual IntegrationsAPPROVED

Reviewed By :Ankita Jodhani 11/22/2021
 Supervised By :Sohil Jodhani 11/22/2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 19 01:41:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR110921CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Nov 10 04:57:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.673	3.852	116.7E6	856947	32.419m	0.291 #
2) SA Decachlor...	10.626	8.860	4129002	3780013	1.236	1.163
Target Compounds						
16) L4 AR-1242-1	5.874	4.933	974.1E6	711.2E6	9962.192	9211.228
17) L4 AR-1242-2	5.897	4.952	1581.6E6	1293.6E6	11585.267	11278.720
18) L4 AR-1242-3	5.961	5.127	709.7E6	601.6E6	8277.714	9742.606
19) L4 AR-1242-4	6.063	5.210	714.1E6	482.3E6	10269.061	7909.764
20) L4 AR-1242-5	6.801	5.729	237.4E6	275.0E6	3114.451	3662.500

AJ
 11/22/21

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.