

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112023\
 Data File : PR064396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 13:35
 Operator : AJ\MA
 Sample : AR1660ICC800
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16604209

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 03:24:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 03:23:33 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.701	3.008	242.1E6	230.1E6	39.706	39.786
2)	SA Decachlor...	9.716	8.387	276.5E6	347.4E6	78.486	80.392
Target Compounds							
3)	L1 AR-1016-1	4.938	4.155	140.6E6	149.2E6	771.700	776.899
4)	L1 AR-1016-2	4.960	4.174	207.1E6	209.6E6	767.673	780.535
5)	L1 AR-1016-3	5.025	4.357	128.3E6	111.9E6	760.592	778.598
6)	L1 AR-1016-4	5.129	4.409	104.6E6	83651338	765.301	778.800
7)	L1 AR-1016-5	5.449	4.632	101.8E6	113.2E6	754.419	772.549
31)	L7 AR-1260-1	6.675	5.748	181.5E6	226.9E6	765.339	780.948
32)	L7 AR-1260-2	6.960	5.956	209.3E6	275.0E6	765.919	781.820
33)	L7 AR-1260-3	7.353	6.118	153.1E6	262.1E6	771.033	791.131
34)	L7 AR-1260-4	7.598	6.633	171.3E6	213.3E6	765.976	792.599
35)	L7 AR-1260-5	7.940	6.901	329.2E6	508.5E6	796.109	812.906

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112023\
Data File : PR064396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 13:35
Operator : AJ\MA
Sample : AR1660ICC800
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16604209

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 03:24:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 03:23:33 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

