

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112023\
 Data File : PR064418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Nov 2023 19:02
 Operator : AJ\MA
 Sample : AR1254ICC100
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12541231

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 21 05:08:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 05:06:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.702	3.008	30562190	29873927	5.168	5.134
2) SA Decachlor...	9.716	8.387	38559173	46773632	10.662	10.354
Target Compounds						
26) L6 AR-1254-1	5.858	5.013	23529170	34321992	115.855	109.484
27) L6 AR-1254-2	6.098	5.175	34872224	30117230	112.811	110.493
28) L6 AR-1254-3	6.494	5.606	34920643	45488301	110.223	106.091
29) L6 AR-1254-4	6.808	5.857	24090576	28475555	110.728	106.931
30) L6 AR-1254-5	7.268	6.310	25768820	41267678	109.114	106.788

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112023\
Data File : PR064418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Nov 2023 19:02
Operator : AJ\MA
Sample : AR1254ICC100
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12541231

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 21 05:08:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 05:06:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

