

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112124\
 Data File : PR069291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2024 15:44
 Operator : AJ\MA
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 04:31:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	0.000	2.982	0	248241	N.D.	0.040 #
2)	SA Decachlor...	0.000	8.261	0	257975	N.D.	0.068 #
Target Compounds							
8)	L2 AR-1221-1	0.000	3.228	0	397646	N.D.	5.308 #
9)	L2 AR-1221-2	0.000	3.248	0	66747	N.D.	1.261 #
10)	L2 AR-1221-3	0.000	3.340	0	151712	N.D.	0.904 #
11)	L3 AR-1232-1	0.000	3.340	0	151712	N.D.	1.131 #
14)	L3 AR-1232-4	0.000	4.413	0	2097710	N.D.	35.238 #
19)	L4 AR-1242-4	0.000	4.413	0	2097710	N.D.	17.313 #
20)	L4 AR-1242-5	0.000	4.915	0	937579	N.D.	6.501 #
23)	L5 AR-1248-3	0.000	4.413	0	2097710	N.D.	11.990 #
25)	L5 AR-1248-5	0.000	4.955	0	5616078	N.D.	28.560 #
26)	L6 AR-1254-1	0.000	4.915	0	937579	N.D.	3.142 #
27)	L6 AR-1254-2	0.000	5.116	0	4051887	N.D.	15.821 #
28)	L6 AR-1254-3	0.000	5.514	0	2931178	N.D.	7.419 #
29)	L6 AR-1254-4	0.000	5.782	0	2240187	N.D.	9.148 #
30)	L6 AR-1254-5	0.000	6.208	0	206594	N.D.	0.592 #
31)	L7 AR-1260-1	0.000	5.657	0	663154	N.D.	2.618 #
32)	L7 AR-1260-2	0.000	5.885	0	1704173	N.D.	5.750 #
33)	L7 AR-1260-3	0.000	6.029	0	458543	N.D.	1.618 #
34)	L7 AR-1260-4	0.000	6.528	0	1370072	N.D.	5.673 #
35)	L7 AR-1260-5	0.000	6.796	0	674673	N.D.	1.183 #
36)	L8 AR-1262-1	0.000	6.254	0	924621	N.D.	2.592 #
37)	L8 AR-1262-2	0.000	6.796	0	674673	N.D.	1.063 #
38)	L8 AR-1262-3	0.000	7.169f	0	776040	N.D.	3.104 #
39)	L8 AR-1262-4	0.000	7.169	0	776040	N.D.	1.634 #
40)	L8 AR-1262-5	0.000	7.675	0	192607	N.D.	0.883 #
41)	L9 AR-1268-1	0.000	7.169f	0	776040	N.D.	1.057 #
42)	L9 AR-1268-2	0.000	7.169	0	776040	N.D.	1.134 #
44)	L9 AR-1268-4	0.000	7.675	0	192607	N.D.	0.779 #
45)	L9 AR-1268-5	0.000	8.008	0	470910	N.D.	0.257 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112124\
Data File : PR069291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2024 15:44
Operator : AJ\MA
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 04:31:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.859 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#8 AR-1221-1

R.T.: 3.228 min
Delta R.T.: 0.034 min
Response: 397646
Conc: 5.31 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 3.947 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 3.248 min
Delta R.T.: -0.031 min
Response: 66747
Conc: 1.26 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.025 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#10 AR-1221-3

R.T.: 3.340 min
Delta R.T.: -0.017 min
Response: 151712
Conc: 0.90 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 4.026 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 3.340 min
Delta R.T.: -0.017 min
Response: 151712
Conc: 1.13 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.047 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#14 AR-1232-4

R.T.: 4.413 min
Delta R.T.: 0.031 min
Response: 2097710
Conc: 35.24 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T.: 5.046 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 4.413 min
Delta R.T.: 0.031 min
Response: 2097710
Conc: 17.31 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 5.834 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#20 AR-1242-5

R.T.: 4.915 min
Delta R.T.: -0.016 min
Response: 937579
Conc: 6.50 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T.: 5.361 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 4.413 min
Delta R.T.: 0.032 min
Response: 2097710
Conc: 11.99 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.834 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 4.955 min
Delta R.T.: -0.019 min
Response: 5616078
Conc: 28.56 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 5.766 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 4.915 min
Delta R.T.: -0.016 min
Response: 937579
Conc: 3.14 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.002 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#27 AR-1254-2

R.T.: 5.116 min
Delta R.T.: 0.025 min
Response: 4051887
Conc: 15.82 ng/ml




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T.: 6.394 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 5.514 min
Delta R.T.: -0.001 min
Response: 2931178
Conc: 7.42 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.705 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#29 AR-1254-4

R.T.: 5.782 min
Delta R.T.: 0.019 min
Response: 2240187
Conc: 9.15 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T.: 7.161 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 6.208 min
Delta R.T.: -0.002 min
Response: 206594
Conc: 0.59 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T.: 6.574 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#31 AR-1260-1

R.T.: 5.657 min
Delta R.T.: 0.000 min
Response: 663154
Conc: 2.62 ng/ml




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 6.856 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 5.885 min
Delta R.T.: 0.022 min
Response: 1704173
Conc: 5.75 ng/ml



#33 AR-1260-3


R.T.: 0.000 min
Exp R.T.: 7.247 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#33 AR-1260-3

R.T.: 6.029 min
Delta R.T.: 0.007 min
Response: 458543
Conc: 1.62 ng/ml




#34 AR-1260-4

R.T.: 0.000 min
Exp R.T.: 7.489 min
Response: 0
Conc: N.D.



#34 AR-1260-4


R.T.: 6.528 min
Delta R.T.: -0.001 min
Response: 1370072
Conc: 5.67 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.829 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#35 AR-1260-5

R.T.: 6.796 min
Delta R.T.: 0.000 min
Response: 674673
Conc: 1.18 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.247 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.254 min
Delta R.T.: -0.001 min
Response: 924621
Conc: 2.59 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.828 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#37 AR-1262-2

R.T.: 6.796 min
Delta R.T.: 0.000 min
Response: 674673
Conc: 1.06 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.134 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.169 min
Delta R.T.: 0.067 min
Response: 776040
Conc: 3.10 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.218 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#39 AR-1262-4

R.T.: 7.169 min
Delta R.T.: 0.000 min
Response: 776040
Conc: 1.63 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 8.841 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 7.675 min
Delta R.T.: -0.033 min
Response: 192607
Conc: 0.88 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.130 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#41 AR-1268-1

R.T.: 7.169 min
Delta R.T.: 0.067 min
Response: 776040
Conc: 1.06 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 8.222 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.169 min
Delta R.T.: -0.003 min
Response: 776040
Conc: 1.13 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.843 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#44 AR-1268-4

R.T.: 7.675 min
Delta R.T.: -0.034 min
Response: 192607
Conc: 0.78 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.226 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 8.008 min
Delta R.T.: 0.002 min
Response: 470910
Conc: 0.26 ng/ml