

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
 Data File : PR058289.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 16:48
 Operator : AJ\MA
 Sample : N5647-06MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXG83MSD

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/23/2022
 Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 06:06:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 17 05:36:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)

R.T.	Response	Conc
4.93	40894225	535.38
4.95	55760262	538.66
5.01	34997922	521.93
5.12	35830855	647.19
5.44	28425597	501.27

(3) AR-1016-1 #2 (L1)

R.T.	Response	Conc
4.03	59500235	569.19
4.05	85705862	601.89
4.23	42581362	556.28
4.28	31755356	538.63
4.49	44755479	567.15

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
Data File : PR058289.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 16:48
Operator : AJ\MA
Sample : N5647-06MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
EXG83MSD

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 06:06:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 17 05:36:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)

R.T.	Response	Conc
4.93	40894225	535.38
4.95	55760262	538.66
5.01	34997922	521.93
5.12	30373751	548.62
5.44	28425597	501.27

(3) AR-1016-1 #2 (L1)

R.T.	Response	Conc
4.03	59500235	569.19
4.05	85705862	601.89
4.23	42581362	556.28
4.28	31755356	538.63
4.49	44755479	567.15

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
Data File : PR058289.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 16:48
Operator : AJ\MA
Sample : N5647-06MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
EXG83MSD

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 06:06:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 17 05:36:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.66	53539565	448.72
6.94	85020366	575.49
7.34	40880070	372.07
7.58	51381591	402.05
7.92	97013881	390.42

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.58	100678222	620.10
5.78	99302425	498.11
5.94	98324482	505.46
6.45	65842556	415.54
6.71	158178547	417.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
 Data File : PR058289.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 16:48
 Operator : AJ\MA
 Sample : N5647-06MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 EXG83MSD

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 11/23/2022
 Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 06:06:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 17 05:36:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.66	53539565	448.72
6.94	85020366	575.49
7.34	40880070	372.07
7.58	51381591	402.05
7.92	97013881	390.42

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.58	100678222	620.10
5.78	99302425	498.11
5.94	98324482	505.46
6.45	65842556	415.54
6.71	164921267	434.96