

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
 Data File : PR058289.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Nov 2022 16:48
 Operator : AJ\MA
 Sample : N5647-06MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EXG83MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/23/2022
 Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 23 06:06:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 17 05:36:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.694	2.916	39304319	51164795	16.930	16.580
2) SA Decachlor...	9.675	8.162	58105024	76166057	26.951	27.315
Target Compounds						
3) L1 AR-1016-1	4.927	4.030	40894225	59500235	535.377	569.189
4) L1 AR-1016-2	4.949	4.048	55760262	85705862	538.663	601.892
5) L1 AR-1016-3	5.014	4.226	34997922	42581362	521.933	556.282
6) L1 AR-1016-4	5.117	4.276	30373751	31755356	548.619m	538.634
7) L1 AR-1016-5	5.436	4.493	28425597	44755479	501.271	567.146
31) L7 AR-1260-1	6.660	5.581	53539565	100.7E6	448.723	620.097 #
32) L7 AR-1260-2	6.945	5.785	85020366	99302425	575.493	498.109
33) L7 AR-1260-3	7.339	5.943	40880070	98324482	372.067	505.459 #
34) L7 AR-1260-4	7.583	6.449	51381591	65842556	402.051	415.539
35) L7 AR-1260-5	7.924	6.713	97013881	164.9E6	390.422	434.955m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112222\
Data File : PR058289.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Nov 2022 16:48
Operator : AJ\MA
Sample : N5647-06MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EXG83MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/23/2022
Supervised By :Ankita Jodhani 11/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 23 06:06:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 17 05:36:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

