



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112423\
Data File : PR064561.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2023 13:39
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 21:38:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 05:53:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit



(26) AR-1254-1 (L6)
R.T. Response Conc
5.87 82914404 408.32
6.11 123715781 402.72
6.51 123230739 392.42
6.83 78214388 362.30
7.29 91609866 392.01

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.01 121016210 390.81
5.17 103808180 386.61
5.61 167999549 396.15
5.86 102791614 389.12
6.31 145911632 386.35

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112423\
Data File : PR064561.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2023 13:39
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 21:38:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 05:53:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)
R.T. Response Conc
5.87 82914404 408.32
6.11 123715781 402.72
6.51 123230739 392.42
6.83 86334461 399.91
7.29 91609866 392.01

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.01 121016210 390.81
5.17 103808180 386.61
5.61 167999549 396.15
5.86 102791614 389.12
6.31 145911632 386.35

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112423\
 Data File : PR064561.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Nov 2023 13:39
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 24 21:38:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 05:53:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.713	3.008	120.1E6	116.8E6	19.711	20.532
2) SA Decachlor...	9.743	8.394	126.9E6	171.3E6	35.949	39.569
Target Compounds						
26) L6 AR-1254-1	5.874	5.012	82914404	121.0E6	408.318	390.813
27) L6 AR-1254-2	6.114	5.175	123.7E6	103.8E6	402.720	386.608
28) L6 AR-1254-3	6.509	5.606	123.2E6	168.0E6	392.418	396.147
29) L6 AR-1254-4	6.825	5.857	86334461	102.8E6	399.912m	389.124
30) L6 AR-1254-5	7.286	6.311	91609866	145.9E6	392.008	386.353

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112423\
Data File : PR064561.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Nov 2023 13:39
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 24 21:38:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 05:53:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

