

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112519\
 Data File : PR043445.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Nov 2019 10:48
 Operator : SM\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 25 11:06:00 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Nov 25 11:05:43 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.708	3.966	216.5E6	692.2E6	50.000	50.000
2) SA Decachlor...	10.619	9.059	305.5E6	832.4E6	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.917	4.178	22162309	76899443	500.000	500.000
9) L2 AR-1221-2	5.005	4.265	16681442	57825936	500.000	500.000
10) L2 AR-1221-3	5.083	4.342	53200364	182.5E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112519\
Data File : PR043445.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Nov 2019 10:48
Operator : SM\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 25 11:06:00 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Nov 25 11:05:43 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.708 min
Delta R.T.: 0.000 min
Response: 216513801
Conc: 50.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.966 min
Delta R.T.: 0.000 min
Response: 692211415
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.619 min
Delta R.T.: 0.000 min
Response: 305530904
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.059 min
Delta R.T.: 0.000 min
Response: 832403011
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 22162309
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.178 min
Delta R.T.: 0.000 min
Response: 76899443
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 5.005 min
Delta R.T.: 0.000 min
Response: 16681442
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.265 min
Delta R.T.: 0.000 min
Response: 57825936
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 5.083 min
Delta R.T.: 0.000 min
Response: 53200364
Conc: 500.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.342 min
Delta R.T.: 0.000 min
Response: 182509792
Conc: 500.00 ng/ml