

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112822\
 Data File : PR058392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Nov 2022 23:14
 Operator : AJ\MA
 Sample : N5721-08MSD
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EXGB7MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/29/2022
 Supervised By :Ankita Jodhani 11/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 29 01:28:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR111622CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Nov 17 05:36:05 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.695	2.916	48839968	64361595	21.038	20.857
2)	SA Decachlor...	9.678	8.160	76785579	102.5E6	35.615	36.763
Target Compounds							
3)	L1 AR-1016-1	4.930	4.029	55510158	76486343	726.725	731.681
4)	L1 AR-1016-2	4.950	4.047	77005182	111.8E6	743.896	785.365
5)	L1 AR-1016-3	5.016	4.225	44505248	58808803	663.718	768.277
6)	L1 AR-1016-4	5.119	4.277	43507312	60198811	785.841	1021.091 #
7)	L1 AR-1016-5	5.431	4.494	43873923	302.5E6	773.695m	3833.810m#
31)	L7 AR-1260-1	6.663	5.582	76271043	204.1E6	639.239	1257.265 #
32)	L7 AR-1260-2	6.930	5.786	268.5E6	102.0E6	1817.557	511.398 #
33)	L7 AR-1260-3	7.342	5.941	52740055	136.8E6	480.010	703.342 #
34)	L7 AR-1260-4	7.585	6.447	59547408	107.9E6	465.947	681.026 #
35)	L7 AR-1260-5	7.926	6.712	118.9E6	203.2E6	478.593	535.865

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112822\
Data File : PR058392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Nov 2022 23:14
Operator : AJ\MA
Sample : N5721-08MSD
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EXGB7MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/29/2022
Supervised By :Ankita Jodhani 11/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 29 01:28:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR11622CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Nov 17 05:36:05 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

