

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR113018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Dec 01 00:24:53 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034166.D 1600 =PR034168.D 800 =PR034167.D
 200 =PR034165.D 100 =PR034164.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.374	1.546	1.448	1.396	1.468	1.446	E6	4.64
2) SA	Decachlorobiphenyl	1.367	1.294	1.324	1.491	1.517	1.399	E6	7.16
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	6.427	6.650	6.149	6.787	7.170	6.637	E4	5.78
27) L6	AR-1254-2	0.995	1.038	0.959	1.060	1.105	1.031	E5	5.47
28) L6	AR-1254-3	1.071	1.142	1.042	1.138	1.171	1.113	E5	4.86
29) L6	AR-1254-4	8.107	8.504	7.860	8.765	8.904	8.428	E4	5.21
30) L6	AR-1254-5	8.533	9.156	8.420	9.195	9.273	8.916	E4	4.54
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR034166.D 1600 =PR034168.D 800 =PR034167.D
 200 =PR034165.D 100 =PR034164.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.793	3.186	2.965	2.824	2.936	2.941	E6	5.27
2) SA	Decachlorobiphenyl	3.321	3.306	3.336	3.539	3.588	3.418	E6	3.93
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR113018CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Dec 01 00:24:53 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034166.D 1600 =PR034168.D 800 =PR034167.D
 200 =PR034165.D 100 =PR034164.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.017	2.195	1.990	2.108	2.216	2.105 E5	4.85
27) L6	AR-1254-2	1.762	1.868	1.721	1.864	1.963	1.836 E5	5.22
28) L6	AR-1254-3	3.014	3.273	2.979	3.137	3.240	3.129 E5	4.20
29) L6	AR-1254-4	1.933	2.083	1.904	2.030	2.095	2.009 E5	4.31
30) L6	AR-1254-5	2.698	2.896	2.640	2.827	2.873	2.787 E5	4.02
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range