

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR113020\
 Data File : PR048764.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Nov 2020 17:56
 Operator : DD\AJ
 Sample : AR1660ICC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 01 03:44:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 03:43:59 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.692	3.845	53548566	85882243	20.000	20.000
2) SA Decachlor...	10.603	8.910	104.9E6	338.8E6	40.000	40.000
Target Compounds						
3) L1 AR-1016-1	5.869	4.931	39107326	28746885	400.000	400.000
4) L1 AR-1016-2	5.892	4.949	56131081	63374793	400.000	400.000
5) L1 AR-1016-3	5.955	5.126	34335496	34218963	400.000	400.000
6) L1 AR-1016-4	6.055	5.167	28213881	15123214	400.000	400.000
7) L1 AR-1016-5	6.350	5.382	28243732	22532683	400.000	400.000
31) L7 AR-1260-1	7.478	6.417	49259755	51708029	400.000	400.000
32) L7 AR-1260-2	7.735	6.606	60677872	185.7E6	400.000	400.000
33) L7 AR-1260-3	8.096	6.760	47043739	111.4E6	400.000	400.000
34) L7 AR-1260-4	8.335	7.235	53881848	135.3E6	400.000	400.000
35) L7 AR-1260-5	8.672	7.478	107.8E6	562.3E6	400.000	400.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR113020\
Data File : PR048764.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Nov 2020 17:56
Operator : DD\AJ
Sample : AR1660ICC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 01 03:44:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 01 03:43:59 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 5.869 min
Delta R.T.: 0.000 min
Response: 39107326
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#3 AR-1016-1

R.T.: 4.931 min
Delta R.T.: 0.000 min
Response: 28746885
Conc: 400.00 ng/ml




#4 AR-1016-2

R.T.: 5.892 min
Delta R.T.: 0.000 min
Response: 56131081
Conc: 400.00 ng/ml



#4 AR-1016-2


R.T.: 4.949 min
Delta R.T.: 0.000 min
Response: 63374793
Conc: 400.00 ng/ml



#5 AR-1016-3


R.T.: 5.955 min
Delta R.T.: 0.000 min
Response: 34335496
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#5 AR-1016-3

R.T.: 5.126 min
Delta R.T.: 0.000 min
Response: 34218963
Conc: 400.00 ng/ml




#6 AR-1016-4

R.T.: 6.055 min
Delta R.T.: 0.000 min
Response: 28213881
Conc: 400.00 ng/ml



#6 AR-1016-4


R.T.: 5.167 min
Delta R.T.: 0.000 min
Response: 15123214
Conc: 400.00 ng/ml



#7 AR-1016-5


R.T.: 6.350 min
Delta R.T.: 0.000 min
Response: 28243732
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#7 AR-1016-5

R.T.: 5.382 min
Delta R.T.: 0.000 min
Response: 22532683
Conc: 400.00 ng/ml




#31 AR-1260-1

R.T.: 7.478 min
Delta R.T.: 0.000 min
Response: 49259755
Conc: 400.00 ng/ml



#31 AR-1260-1


R.T.: 6.417 min
Delta R.T.: 0.000 min
Response: 51708029
Conc: 400.00 ng/ml



#32 AR-1260-2


R.T.: 7.735 min
Delta R.T.: 0.000 min
Response: 60677872
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#32 AR-1260-2

R.T.: 6.606 min
Delta R.T.: 0.000 min
Response: 185748662
Conc: 400.00 ng/ml




#33 AR-1260-3

R.T.: 8.096 min
Delta R.T.: 0.000 min
Response: 47043739
Conc: 400.00 ng/ml



#33 AR-1260-3


R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 111355052
Conc: 400.00 ng/ml



#34 AR-1260-4


R.T.: 8.335 min
Delta R.T.: 0.000 min
Response: 53881848
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#34 AR-1260-4

R.T.: 7.235 min
Delta R.T.: 0.000 min
Response: 135344563
Conc: 400.00 ng/ml



#35 AR-1260-5

R.T.: 8.672 min
Delta R.T.: 0.000 min
Response: 107804124
Conc: 400.00 ng/ml



#35 AR-1260-5

R.T.: 7.478 min
Delta R.T.: 0.000 min
Response: 562267287
Conc: 400.00 ng/ml