

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120320\
Data File : PR048825.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Dec 2020 17:15
Operator : DD\AJ
Sample : L4941-02
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
C0AX3

Manual Integrations
APPROVED

mohammad
12/4/2020 12:35:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 04 01:51:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 01 06:11:17 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.73 11026169 100.63
6.95 17052058 102.36
7.32 8078779 46.26
7.60 14634304 108.34
8.02 22856957 159.49

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.74 18183528 151.94
5.89 13745461 111.35
6.29 23487094 93.09
6.52 29763520 96.85
6.94 87152883 199.58

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120320\
 Data File : PR048825.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2020 17:15
 Operator : DD\AJ
 Sample : L4941-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 C0AX3

Manual Integrations
 APPROVED

mohammad
 12/4/2020 12:35:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 01:51:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 #2 (L6)

R.T.	Response	Conc
6.73	9855211	89.95
6.94	15922548	95.58
7.32	8078779	46.26
7.60	12795592	94.73
8.02	21566127	150.48

(26) AR-1254-1 #2 (L6)

R.T.	Response	Conc
5.74	18183528	151.94
5.89	13745461	111.35
6.29	25963374	102.90
6.52	29763520	96.85
6.94	87152883	199.58

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120320\
 Data File : PR048825.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Dec 2020 17:15
 Operator : DD\AJ
 Sample : L4941-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AX3

Manual Integrations
 APPROVED

mohammad
 12/4/2020 12:35:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 04 01:51:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.694	3.849	26513905	48217400	10.033	11.336
2) SA Decachlor...	10.600	8.916	37777305	124.4E6	14.396	14.512
Target Compounds						
26) L6 AR-1254-1	6.727	5.739	9855211	18183528	89.946m	151.939 #
27) L6 AR-1254-2	6.945	5.885	15922548	13745461	95.577m	111.351
28) L6 AR-1254-3	7.319	6.289	8078779	25963374	46.263	102.903m#
29) L6 AR-1254-4	7.601	6.520	12795592	29763520	94.730m	96.854
30) L6 AR-1254-5	8.023	6.938	21566127	87152883	150.479m	199.578 #

AJ
 12/11/20

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.