

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
 Data File : PR043491.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2019 09:04
 Operator : SM\AJ
 Sample : AIBLK51
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 03:53:15 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:51:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.710	3.969	101.1E6	147.8E6	18.181	19.137
2)	SA Decachlor...	10.626	9.064	143.1E6	381.2E6	37.994	38.517
Target Compounds							
6)	L1 AR-1016-4	6.095	0.000	593153	0	4.307	N.D. #
9)	L2 AR-1221-2	5.022	0.000	1270196	0	31.081	N.D. #
10)	L2 AR-1221-3	5.022f	0.000	1270196	0	9.242	N.D. #
11)	L3 AR-1232-1	5.022f	0.000	1270196	0	11.817	N.D. #
14)	L3 AR-1232-4	6.095	0.000	593153	0	10.205	N.D. #
15)	L3 AR-1232-5	6.184	0.000	752070	0	16.844	N.D. #
19)	L4 AR-1242-4	6.095	0.000	593153	0	5.167	N.D. #
20)	L4 AR-1242-5	6.748f	0.000	3029013	0	23.813	N.D. #
22)	L5 AR-1248-2	6.184	0.000	752070	0	4.616	N.D. #
24)	L5 AR-1248-4	6.748	0.000	3029013	0	14.469	N.D. #
25)	L5 AR-1248-5	6.748f	0.000	3029013	0	15.257	N.D. #
26)	L6 AR-1254-1	6.748	0.000	3029013	0	13.858	N.D. #
27)	L6 AR-1254-2	6.968	0.000	632733	0	1.876	N.D. #
28)	L6 AR-1254-3	7.350	0.000	1144318	0	3.272	N.D. #
29)	L6 AR-1254-4	7.605	6.623	1012837	2011766	4.053	2.794 #
32)	L7 AR-1260-2	7.759	6.694f	457300	1099096	1.539	1.134 #
35)	L7 AR-1260-5	8.667	0.000	567916	0	1.178	N.D. #
37)	L8 AR-1262-2	8.667	0.000	567916	0	0.997	N.D. #
40)	L8 AR-1262-5	9.800	0.000	2992049	0	15.889	N.D. #
43)	L9 AR-1268-3	9.363	8.169	1354536	3386724	2.877	2.200
44)	L9 AR-1268-4	9.800	0.000	2992049	0	14.644	N.D. #
45)	L9 AR-1268-5	10.264	8.789	1496602	6264394	0.957	1.481 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
Data File : PR043491.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2019 09:04
Operator : SM\AJ
Sample : AIBLK51
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 03:53:15 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:51:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#6 AR-1016-4


R.T.: 6.095 min
Delta R.T.: 0.021 min
Response: 593153
Conc: 4.31 ng/ml

Instrument :
ECD_R
ClientSampleId :




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.300 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.022 min
Delta R.T.: 0.018 min
Response: 1270196
Conc: 31.08 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.268 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 5.022 min
Delta R.T.: -0.061 min
Response: 1270196
Conc: 9.24 ng/ml

Instrument :
ECD_R
ClientSampleId :




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 4.344 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 5.022 min
Delta R.T.: -0.062 min
Response: 1270196
Conc: 11.82 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 4.344 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 6.095 min
Delta R.T.: 0.023 min
Response: 593153
Conc: 10.20 ng/ml

Instrument :
ECD_R
ClientSampleId :




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 5.343 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 6.184 min
Delta R.T.: 0.024 min
Response: 752070
Conc: 16.84 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.517 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 6.095 min
Delta R.T.: 0.022 min
Response: 593153
Conc: 5.17 ng/ml

Instrument :
ECD_R
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.343 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.748 min
Delta R.T.: -0.063 min
Response: 3029013
Conc: 23.81 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T. : 5.868 min
Response: 0
Conc: N.D.



#22 AR-1248-2


R.T.: 6.184 min
Delta R.T.: 0.024 min
Response: 752070
Conc: 4.62 ng/ml

Instrument :
ECD_R
ClientSampleId :




#22 AR-1248-2

R.T.: 0.000 min
Exp R.T. : 5.300 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.748 min
Delta R.T.: -0.024 min
Response: 3029013
Conc: 14.47 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.516 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.748 min
Delta R.T.: -0.063 min
Response: 3029013
Conc: 15.26 ng/ml

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.911 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.748 min
Delta R.T.: 0.004 min
Response: 3029013
Conc: 13.86 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.868 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 6.968 min
Delta R.T.: 0.006 min
Response: 632733
Conc: 1.88 ng/ml

Instrument :
ECD_R
ClientSampleId :




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 6.016 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.350 min
Delta R.T.: 0.018 min
Response: 1144318
Conc: 3.27 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.422 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.605 min
Delta R.T.: -0.013 min
Response: 1012837
Conc: 4.05 ng/ml

Instrument :
ECD_R
ClientSampleId :




#29 AR-1254-4

R.T.: 6.623 min
Delta R.T.: -0.026 min
Response: 2011766
Conc: 2.79 ng/ml




#31 AR-1260-1

R.T.: 7.540 min
Delta R.T.: 0.045 min
Response: 38346
Conc: 0.15 ng/ml



#31 AR-1260-1


R.T.: 6.573 min
Delta R.T.: 0.020 min
Response: -324284
Conc: N.D.



#32 AR-1260-2


R.T.: 7.759 min
Delta R.T.: 0.008 min
Response: 457300
Conc: 1.54 ng/ml

Instrument :
ECD_R
ClientSampleId :




#32 AR-1260-2

R.T.: 6.694 min
Delta R.T.: -0.046 min
Response: 1099096
Conc: 1.13 ng/ml




#35 AR-1260-5

R.T.: 8.667 min
Delta R.T.: -0.022 min
Response: 567916
Conc: 1.18 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.609 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.667 min
Delta R.T.: -0.019 min
Response: 567916
Conc: 1.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.607 min
Response: 0
Conc: N.D.




#40 AR-1262-5

R.T.: 9.800 min
Delta R.T.: -0.017 min
Response: 2992049
Conc: 15.89 ng/ml



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T. : 8.472 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 9.363 min
Delta R.T.: 0.003 min
Response: 1354536
Conc: 2.88 ng/ml

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 8.169 min
Delta R.T.: 0.000 min
Response: 3386724
Conc: 2.20 ng/ml




#44 AR-1268-4

R.T.: 9.800 min
Delta R.T.: -0.018 min
Response: 2992049
Conc: 14.64 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T. : 8.472 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 10.264 min
Delta R.T.: 0.003 min
Response: 1496602
Conc: 0.96 ng/ml

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 8.789 min
Delta R.T.: 0.002 min
Response: 6264394
Conc: 1.48 ng/ml