

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
 Data File : PR043518.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2019 15:35
 Operator : SM\AJ
 Sample : AR1254ICC200
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 03:19:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:18:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.708	3.967	61137397	90213704	9.867	10.233
2) SA Decachlor...	10.626	9.065	85695494	215.0E6	21.076	20.724
Target Compounds						
26) L6 AR-1254-1	6.745	5.870	46006617	152.5E6	212.268	216.004
27) L6 AR-1254-2	6.962	6.016	71739156	138.1E6	217.993	218.949
28) L6 AR-1254-3	7.332	6.423	74655399	218.3E6	221.615	222.838
29) L6 AR-1254-4	7.618	6.650	53941612	156.8E6	225.816	225.932
30) L6 AR-1254-5	8.038	7.070	55928762	185.8E6	228.050	225.056



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
Data File : PR043518.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2019 15:35
Operator : SM\AJ
Sample : AR1254ICC200
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 03:19:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:18:36 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 46006617
Conc: 212.27 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 152519007
Conc: 216.00 ng/ml




#27 AR-1254-2

R.T.: 6.962 min
Delta R.T.: 0.000 min
Response: 71739156
Conc: 217.99 ng/ml



#27 AR-1254-2


R.T.: 6.016 min
Delta R.T.: 0.000 min
Response: 138146485
Conc: 218.95 ng/ml



#28 AR-1254-3


R.T.: 7.332 min
Delta R.T.: 0.000 min
Response: 74655399
Conc: 221.62 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 6.423 min
Delta R.T.: 0.000 min
Response: 218266409
Conc: 222.84 ng/ml




#29 AR-1254-4

R.T.: 7.618 min
Delta R.T.: 0.000 min
Response: 53941612
Conc: 225.82 ng/ml



#29 AR-1254-4


R.T.: 6.650 min
Delta R.T.: 0.000 min
Response: 156827843
Conc: 225.93 ng/ml



#30 AR-1254-5

R.T.: 8.038 min
Delta R.T.: 0.000 min
Response: 55928762
Conc: 228.05 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 7.070 min
Delta R.T.: 0.000 min
Response: 185775950
Conc: 225.06 ng/ml