

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
 Data File : PR043521.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Dec 2019 16:18
 Operator : SM\AJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 03:20:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:18:36 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.709	3.968	507.7E6	662.6E6	81.929	75.160
2) SA Decachlor...	10.623	9.064	611.0E6	1561.1E6	150.281	150.502
Target Compounds						
26) L6 AR-1254-1	6.745	5.869	317.6E6	1001.7E6	1465.331	1418.639
27) L6 AR-1254-2	6.962	6.016	486.1E6	888.1E6	1476.958	1407.630
28) L6 AR-1254-3	7.332	6.422	506.9E6	1414.7E6	1504.653	1444.300
29) L6 AR-1254-4	7.617	6.650	355.3E6	1002.1E6	1487.326	1443.595
30) L6 AR-1254-5	8.039	7.069	366.1E6	1192.8E6	1492.665	1445.031

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120419\
Data File : PR043521.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Dec 2019 16:18
Operator : SM\AJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1254501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 03:20:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:18:36 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

