

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120619\
 Data File : PR043658.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 19:31
 Operator : SM\AJ
 Sample : K6160-03
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P001-CK45-0612-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 05:10:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:51:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.707	3.966	84635357	115.7E6	15.224	14.978
2)	SA Decachlor...	10.616	9.057	112.5E6	274.2E6	29.865	27.705
Target Compounds							
3)	L1 AR-1016-1	5.883	5.065	2345842	2069727	12.115	5.077 #
4)	L1 AR-1016-2	5.913	5.089	4652894	203119	16.993	0.364 #
5)	L1 AR-1016-3	5.913f	0.000	4652894	0	28.253	N.D. #
6)	L1 AR-1016-4	6.041	0.000	1725251	0	12.527	N.D. #
7)	L1 AR-1016-5	6.363	0.000	4825135	0	34.919	N.D. #
8)	L2 AR-1221-1	4.925	4.227f	175472	1706935	3.059	20.021 #
9)	L2 AR-1221-2	5.019	4.227f	2853494	1706935	69.825	27.883 #
10)	L2 AR-1221-3	5.094	0.000	3071844	0	22.350	N.D. #
11)	L3 AR-1232-1	5.094	0.000	3071844	0	28.578	N.D. #
12)	L3 AR-1232-2	5.550f	5.089	3739636	203119	67.568	0.870 #
13)	L3 AR-1232-3	5.913	0.000	4652894	0	41.033	N.D. #
14)	L3 AR-1232-4	6.041	0.000	1725251	0	29.682	N.D. #
15)	L3 AR-1232-5	6.170	0.000	3064993	0	68.646	N.D. #
16)	L4 AR-1242-1	5.883	5.065	2345842	2069727	14.431	6.260 #
17)	L4 AR-1242-2	5.913	5.089	4652894	203119	20.397	0.444 #
18)	L4 AR-1242-3	5.913f	0.000	4652894	0	33.827	N.D. #
19)	L4 AR-1242-4	6.041	0.000	1725251	0	15.030	N.D. #
20)	L4 AR-1242-5	6.806	5.869	3801708	5274164	29.888	15.472 #
21)	L5 AR-1248-1	5.883	5.065	2345842	2069727	19.764	8.458 #
22)	L5 AR-1248-2	6.170	0.000	3064993	0	18.813	N.D. #
23)	L5 AR-1248-3	6.363	0.000	4825135	0	26.574	N.D. #
24)	L5 AR-1248-4	6.785	0.000	4390928	0	20.975	N.D. #
25)	L5 AR-1248-5	6.806	5.869f	3801708	5274164	19.149	11.867 #
26)	L6 AR-1254-1	6.738	5.869	5278325	5274164	24.149	7.403 #
27)	L6 AR-1254-2	6.958	6.015	4402902	7050191	13.057	11.022
28)	L6 AR-1254-3	7.329	6.420	1980990	19219784	5.664	18.987 #
29)	L6 AR-1254-4	7.614	6.648	2711019	38995583	10.849	54.155 #
30)	L6 AR-1254-5	8.034	7.067	8028533	42320549	31.155	49.679 #
31)	L7 AR-1260-1	7.491	6.554	3601564	30294644	14.453	44.716 #
32)	L7 AR-1260-2	7.745	6.736	6305985	68464408	21.225	70.665 #
33)	L7 AR-1260-3	8.106	6.892	1998423	44392639	9.020	56.845 #
34)	L7 AR-1260-4	8.338	7.363	6127937	11049417	24.362	17.417 #
35)	L7 AR-1260-5	8.683	7.604	3894360	26790132	8.076	15.739 #
36)	L8 AR-1262-1	8.162	7.067	1545939	42320549	11.613	85.584 #
37)	L8 AR-1262-2	8.683	7.604	3894360	26790132	6.837	14.081 #
38)	L8 AR-1262-3	9.034	7.890	2915551	10283253	8.053	13.481 #
39)	L8 AR-1262-4	9.160f	7.960	9395875	14454264	35.302	11.644 #
40)	L8 AR-1262-5	9.811	8.468	1302505	9296983	6.917	16.312 #
41)	L9 AR-1268-1	9.034	7.890	2915551	10283253	4.556	4.792

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120619\
 Data File : PR043658.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2019 19:31
 Operator : SM\AJ
 Sample : K6160-03
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 P001-CK45-0612-01

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 07 05:10:27 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:51:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.160f	7.960	9395875	14454264	16.307	7.966 #
43) L9	AR-1268-3	9.353	8.163	814032	6261258	1.729	4.067 #
44) L9	AR-1268-4	9.811	8.468	1302505	9296983	6.375	15.161 #
45) L9	AR-1268-5	10.256	8.781	2513234	10496888	1.607	2.481 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120619\
Data File : PR043658.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2019 19:31
Operator : SM\AJ
Sample : K6160-03
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 07 05:10:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:51:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 5.883 min
Delta R.T.: -0.005 min
Response: 2345842
Conc: 12.12 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#3 AR-1016-1

R.T.: 5.065 min
Delta R.T.: 0.003 min
Response: 2069727
Conc: 5.08 ng/ml




#4 AR-1016-2

R.T.: 5.913 min
Delta R.T.: 0.002 min
Response: 4652894
Conc: 16.99 ng/ml



#4 AR-1016-2


R.T.: 5.089 min
Delta R.T.: 0.007 min
Response: 203119
Conc: 0.36 ng/ml



#5 AR-1016-3


R.T.: 5.913 min
Delta R.T.: -0.060 min
Response: 4652894
Conc: 28.25 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 5.260 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 6.041 min
Delta R.T.: -0.033 min
Response: 1725251
Conc: 12.53 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 5.300 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.363 min
Delta R.T.: -0.005 min
Response: 4825135
Conc: 34.92 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.517 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.925 min
Delta R.T.: 0.006 min
Response: 175472
Conc: 3.06 ng/ml



#8 AR-1221-1


R.T.: 4.227 min
Delta R.T.: 0.047 min
Response: 1706935
Conc: 20.02 ng/ml



#9 AR-1221-2


R.T.: 5.019 min
Delta R.T.: 0.014 min
Response: 2853494
Conc: 69.82 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#9 AR-1221-2

R.T.: 4.227 min
Delta R.T.: -0.042 min
Response: 1706935
Conc: 27.88 ng/ml




#10 AR-1221-3

R.T.: 5.094 min
Delta R.T.: 0.010 min
Response: 3071844
Conc: 22.35 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.344 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 5.094 min
Delta R.T.: 0.009 min
Response: 3071844
Conc: 28.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.344 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.550 min
Delta R.T.: -0.069 min
Response: 3739636
Conc: 67.57 ng/ml



#12 AR-1232-2


R.T.: 5.089 min
Delta R.T.: 0.007 min
Response: 203119
Conc: 0.87 ng/ml



#13 AR-1232-3


R.T.: 5.913 min
Delta R.T.: 0.003 min
Response: 4652894
Conc: 41.03 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 5.260 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 6.041 min
Delta R.T.: -0.031 min
Response: 1725251
Conc: 29.68 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 5.343 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 6.170 min
Delta R.T.: 0.010 min
Response: 3064993
Conc: 68.65 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.517 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.883 min
Delta R.T.: -0.004 min
Response: 2345842
Conc: 14.43 ng/ml



#16 AR-1242-1


R.T.: 5.065 min
Delta R.T.: 0.004 min
Response: 2069727
Conc: 6.26 ng/ml



#17 AR-1242-2


R.T.: 5.913 min
Delta R.T.: 0.003 min
Response: 4652894
Conc: 20.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#17 AR-1242-2

R.T.: 5.089 min
Delta R.T.: 0.008 min
Response: 203119
Conc: 0.44 ng/ml




#18 AR-1242-3

R.T.: 5.913 min
Delta R.T.: -0.060 min
Response: 4652894
Conc: 33.83 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T.: 5.260 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 6.041 min
Delta R.T.: -0.032 min
Response: 1725251
Conc: 15.03 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.343 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.806 min
Delta R.T.: -0.005 min
Response: 3801708
Conc: 29.89 ng/ml



#20 AR-1242-5


R.T.: 5.869 min
Delta R.T.: 0.000 min
Response: 5274164
Conc: 15.47 ng/ml



#21 AR-1248-1


R.T.: 5.883 min
Delta R.T.: -0.004 min
Response: 2345842
Conc: 19.76 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#21 AR-1248-1

R.T.: 5.065 min
Delta R.T.: 0.003 min
Response: 2069727
Conc: 8.46 ng/ml




#22 AR-1248-2

R.T.: 6.170 min
Delta R.T.: 0.010 min
Response: 3064993
Conc: 18.81 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.300 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.363 min
Delta R.T.: -0.005 min
Response: 4825135
Conc: 26.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 5.342 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.785 min
Delta R.T.: 0.014 min
Response: 4390928
Conc: 20.98 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.516 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.806 min
Delta R.T.: -0.005 min
Response: 3801708
Conc: 19.15 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#25 AR-1248-5

R.T.: 5.869 min
Delta R.T.: -0.042 min
Response: 5274164
Conc: 11.87 ng/ml




#26 AR-1254-1

R.T.: 6.738 min
Delta R.T.: -0.006 min
Response: 5278325
Conc: 24.15 ng/ml



#26 AR-1254-1


R.T.: 5.869 min
Delta R.T.: 0.000 min
Response: 5274164
Conc: 7.40 ng/ml



#27 AR-1254-2


R.T.: 6.958 min
Delta R.T.: -0.003 min
Response: 4402902
Conc: 13.06 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#27 AR-1254-2

R.T.: 6.015 min
Delta R.T.: 0.000 min
Response: 7050191
Conc: 11.02 ng/ml




#28 AR-1254-3

R.T.: 7.329 min
Delta R.T.: -0.003 min
Response: 1980990
Conc: 5.66 ng/ml



#28 AR-1254-3


R.T.: 6.420 min
Delta R.T.: -0.002 min
Response: 19219784
Conc: 18.99 ng/ml



#29 AR-1254-4


R.T.: 7.614 min
Delta R.T.: -0.005 min
Response: 2711019
Conc: 10.85 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#29 AR-1254-4

R.T.: 6.648 min
Delta R.T.: -0.002 min
Response: 38995583
Conc: 54.15 ng/ml




#30 AR-1254-5

R.T.: 8.034 min
Delta R.T.: -0.004 min
Response: 8028533
Conc: 31.15 ng/ml



#30 AR-1254-5


R.T.: 7.067 min
Delta R.T.: -0.003 min
Response: 42320549
Conc: 49.68 ng/ml



#31 AR-1260-1


R.T.: 7.491 min
Delta R.T.: -0.004 min
Response: 3601564
Conc: 14.45 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#31 AR-1260-1

R.T.: 6.554 min
Delta R.T.: 0.001 min
Response: 30294644
Conc: 44.72 ng/ml




#32 AR-1260-2

R.T.: 7.745 min
Delta R.T.: -0.006 min
Response: 6305985
Conc: 21.23 ng/ml



#32 AR-1260-2


R.T.: 6.736 min
Delta R.T.: -0.003 min
Response: 68464408
Conc: 70.66 ng/ml



#33 AR-1260-3


R.T.: 8.106 min
Delta R.T.: -0.006 min
Response: 1998423
Conc: 9.02 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#33 AR-1260-3

R.T.: 6.892 min
Delta R.T.: -0.004 min
Response: 44392639
Conc: 56.85 ng/ml




#34 AR-1260-4

R.T.: 8.338 min
Delta R.T.: -0.013 min
Response: 6127937
Conc: 24.36 ng/ml



#34 AR-1260-4


R.T.: 7.363 min
Delta R.T.: -0.005 min
Response: 11049417
Conc: 17.42 ng/ml



#35 AR-1260-5


R.T.: 8.683 min
Delta R.T.: -0.005 min
Response: 3894360
Conc: 8.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#35 AR-1260-5

R.T.: 7.604 min
Delta R.T.: -0.005 min
Response: 26790132
Conc: 15.74 ng/ml




#36 AR-1262-1

R.T.: 8.162 min
Delta R.T.: -0.008 min
Response: 1545939
Conc: 11.61 ng/ml



#36 AR-1262-1


R.T.: 7.067 min
Delta R.T.: 0.000 min
Response: 42320549
Conc: 85.58 ng/ml



#37 AR-1262-2


R.T.: 8.683 min
Delta R.T.: -0.003 min
Response: 3894360
Conc: 6.84 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#37 AR-1262-2

R.T.: 7.604 min
Delta R.T.: -0.003 min
Response: 26790132
Conc: 14.08 ng/ml




#38 AR-1262-3

R.T.: 9.034 min
Delta R.T.: 0.017 min
Response: 2915551
Conc: 8.05 ng/ml



#38 AR-1262-3


R.T.: 7.890 min
Delta R.T.: -0.002 min
Response: 10283253
Conc: 13.48 ng/ml



#39 AR-1262-4


R.T.: 9.160 min
Delta R.T.: 0.047 min
Response: 9395875
Conc: 35.30 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#39 AR-1262-4

R.T.: 7.960 min
Delta R.T.: 0.004 min
Response: 14454264
Conc: 11.64 ng/ml




#40 AR-1262-5

R.T.: 9.811 min
Delta R.T.: -0.007 min
Response: 1302505
Conc: 6.92 ng/ml



#40 AR-1262-5


R.T.: 8.468 min
Delta R.T.: -0.004 min
Response: 9296983
Conc: 16.31 ng/ml



#41 AR-1268-1


R.T.: 9.034 min
Delta R.T.: 0.019 min
Response: 2915551
Conc: 4.56 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#41 AR-1268-1

R.T.: 7.890 min
Delta R.T.: -0.003 min
Response: 10283253
Conc: 4.79 ng/ml




#42 AR-1268-2

R.T.: 9.160 min
Delta R.T.: 0.043 min
Response: 9395875
Conc: 16.31 ng/ml



#42 AR-1268-2


R.T.: 7.960 min
Delta R.T.: 0.002 min
Response: 14454264
Conc: 7.97 ng/ml



#43 AR-1268-3


R.T.: 9.353 min
Delta R.T.: -0.006 min
Response: 814032
Conc: 1.73 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01




#43 AR-1268-3

R.T.: 8.163 min
Delta R.T.: -0.006 min
Response: 6261258
Conc: 4.07 ng/ml




#44 AR-1268-4

R.T.: 9.811 min
Delta R.T.: -0.007 min
Response: 1302505
Conc: 6.37 ng/ml



#44 AR-1268-4


R.T.: 8.468 min
Delta R.T.: -0.004 min
Response: 9296983
Conc: 15.16 ng/ml



#45 AR-1268-5

R.T.: 10.256 min
Delta R.T.: -0.005 min
Response: 2513234
Conc: 1.61 ng/ml

Instrument :
ECD_R
ClientSampleId :
P001-CK45-0612-01



#45 AR-1268-5

R.T.: 8.781 min
Delta R.T.: -0.006 min
Response: 10496888
Conc: 2.48 ng/ml