

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120623\
 Data File : PR064599.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2023 12:57
 Operator : AJ\MA
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR16603400

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/07/2023
 Supervised By :Ankita Jodhani 12/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 21:47:51 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Nov 21 05:53:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.703	3.010	120.4E6	109.3E6	19.758	19.216
2) SA Decachlor...	9.723	8.389	136.8E6	160.0E6	38.741	36.942
Target Compounds						
3) L1 AR-1016-1	4.941	4.156	69596746	71028436	381.862	372.208
4) L1 AR-1016-2	4.963	4.175	104.4E6	99668495	389.391	375.759
5) L1 AR-1016-3	5.029	4.360	66376746	53779606	397.447	376.874
6) L1 AR-1016-4	5.132	4.410	52520870	42113584	384.797	394.102
7) L1 AR-1016-5	5.451	4.634	53329080	55183726	402.652	384.234
31) L7 AR-1260-1	6.678	5.749	94297748	106.6E6	394.505	364.808m
32) L7 AR-1260-2	6.965	5.957	106.2E6	125.1E6	388.132	356.795m
33) L7 AR-1260-3	7.358	6.120	79243643	120.1E6	398.779	356.452
34) L7 AR-1260-4	7.604	6.635	87036854	95940275	393.972	357.059
35) L7 AR-1260-5	7.945	6.903	158.9E6	226.2E6	382.857	359.079m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120623\
Data File : PR064599.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2023 12:57
Operator : AJ\MA
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR16603400

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/07/2023
Supervised By :Ankita Jodhani 12/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 21:47:51 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112023CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Nov 21 05:53:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

