

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
 Data File : PR069391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 19:21
 Operator : AJ\MA
 Sample : AIBLK091
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AIBLK114

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 06 23:04:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.637	2.969	18886464	118.5E6	18.055	19.034
2) SA Decachlor...	9.528	8.252	22908213	150.3E6	42.652	39.885
Target Compounds						
6) L1 AR-1016-4	0.000	4.367	0	2189197	N.D.	17.686 #
7) L1 AR-1016-5	0.000	4.517f	0	126488	N.D.	0.836 #
8) L2 AR-1221-1	0.000	3.226	0	509112	N.D.	6.795 #
9) L2 AR-1221-2	0.000	3.273	0	1801107	N.D.	34.028 #
10) L2 AR-1221-3	0.000	3.340	0	226461	N.D.	1.349 #
11) L3 AR-1232-1	0.000	3.340	0	226461	N.D.	1.688 #
14) L3 AR-1232-4	0.000	4.367	0	2189197	N.D.	36.775 #
15) L3 AR-1232-5	0.000	4.517f	0	126488	N.D.	2.003 #
19) L4 AR-1242-4	0.000	4.367	0	2189197	N.D.	18.068 #
20) L4 AR-1242-5	0.000	4.954	0	770312	N.D.	5.341 #
22) L5 AR-1248-2	0.000	4.367	0	2189197	N.D.	12.636 #
23) L5 AR-1248-3	0.000	4.367	0	2189197	N.D.	12.513 #
24) L5 AR-1248-4	0.000	4.517f	0	126488	N.D.	0.616 #
25) L5 AR-1248-5	0.000	4.954	0	770312	N.D.	3.917 #
26) L6 AR-1254-1	0.000	4.954	0	770312	N.D.	2.582 #
27) L6 AR-1254-2	0.000	5.092	0	1157455	N.D.	4.519 #
28) L6 AR-1254-3	0.000	5.495	0	2720708	N.D.	6.886 #
29) L6 AR-1254-4	0.000	5.710f	0	2303776	N.D.	9.407 #
30) L6 AR-1254-5	0.000	6.247f	0	328242	N.D.	0.941 #
31) L7 AR-1260-1	0.000	5.618f	0	2079820	N.D.	8.210 #
32) L7 AR-1260-2	0.000	5.870	0	1514051	N.D.	5.109 #
33) L7 AR-1260-3	0.000	6.028	0	303896	N.D.	1.072 #
35) L7 AR-1260-5	0.000	6.865f	0	1061872	N.D.	1.862 #
36) L8 AR-1262-1	0.000	6.247	0	328242	N.D.	0.920 #
37) L8 AR-1262-2	0.000	6.865f	0	1061872	N.D.	1.673 #
38) L8 AR-1262-3	0.000	7.092	0	289484	N.D.	1.158 #
41) L9 AR-1268-1	0.000	7.092	0	289484	N.D.	0.394 #
43) L9 AR-1268-3	0.000	7.393	0	324379	N.D.	0.565 #
45) L9 AR-1268-5	0.000	8.004	0	1460109	N.D.	0.798 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
Data File : PR069391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 19:21
Operator : AJ\MA
Sample : AIBLK091
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AIBLK114

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 06 23:04:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#6 AR-1016-4

R.T.: 0.000 min
Exp R.T.: 5.047 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114


Time



#6 AR-1016-4

R.T.: 4.367 min
Delta R.T.: 0.027 min
Response: 2189197
Conc: 17.69 ng/ml


Time



#7 AR-1016-5

R.T.: 0.000 min
Exp R.T.: 5.362 min
Response: 0
Conc: N.D.


Time



#7 AR-1016-5

R.T.: 4.517 min
Delta R.T.: -0.042 min
Response: 126488
Conc: 0.84 ng/ml


Time



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 3.859 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#8 AR-1221-1

R.T.: 3.226 min
Delta R.T.: 0.033 min
Response: 509112
Conc: 6.80 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T.: 3.947 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 3.273 min
Delta R.T.: -0.007 min
Response: 1801107
Conc: 34.03 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 4.025 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#10 AR-1221-3

R.T.: 3.340 min
Delta R.T.: -0.017 min
Response: 226461
Conc: 1.35 ng/ml




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T.: 4.026 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 3.340 min
Delta R.T.: -0.017 min
Response: 226461
Conc: 1.69 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T.: 5.047 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#14 AR-1232-4

R.T.: 4.367 min
Delta R.T.: -0.015 min
Response: 2189197
Conc: 36.77 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T.: 5.150 min
Response: 0
Conc: N.D.



#15 AR-1232-5

R.T.: 4.517 min
Delta R.T.: -0.042 min
Response: 126488
Conc: 2.00 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T.: 5.046 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114


Time



#19 AR-1242-4

R.T.: 4.367 min
Delta R.T.: -0.015 min
Response: 2189197
Conc: 18.07 ng/ml


Time



#20 AR-1242-5

R.T.: 0.000 min
Exp R.T.: 5.834 min
Response: 0
Conc: N.D.


Time



#20 AR-1242-5

R.T.: 4.954 min
Delta R.T.: 0.022 min
Response: 770312
Conc: 5.34 ng/ml


Time



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T.: 5.150 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#22 AR-1248-2

R.T.: 4.367 min
Delta R.T.: 0.027 min
Response: 2189197
Conc: 12.64 ng/ml




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T.: 5.361 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 4.367 min
Delta R.T.: -0.015 min
Response: 2189197
Conc: 12.51 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.795 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#24 AR-1248-4

R.T.: 4.517 min
Delta R.T.: -0.042 min
Response: 126488
Conc: 0.62 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T.: 5.834 min
Response: 0
Conc: N.D.



#25 AR-1248-5

R.T.: 4.954 min
Delta R.T.: -0.020 min
Response: 770312
Conc: 3.92 ng/ml




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T.: 5.766 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114


Time



#26 AR-1254-1

R.T.: 4.954 min
Delta R.T.: 0.022 min
Response: 770312
Conc: 2.58 ng/ml


Time



#27 AR-1254-2

R.T.: 0.000 min
Exp R.T.: 6.002 min
Response: 0
Conc: N.D.


Time



#27 AR-1254-2

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 1157455
Conc: 4.52 ng/ml


Time



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.394 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#28 AR-1254-3

R.T.: 5.495 min
Delta R.T.: -0.021 min
Response: 2720708
Conc: 6.89 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T.: 6.705 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 5.710 min
Delta R.T.: -0.053 min
Response: 2303776
Conc: 9.41 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 7.161 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#30 AR-1254-5

R.T.: 6.247 min
Delta R.T.: 0.038 min
Response: 328242
Conc: 0.94 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T.: 6.574 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 5.618 min
Delta R.T.: -0.039 min
Response: 2079820
Conc: 8.21 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 6.856 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#32 AR-1260-2

R.T.: 5.870 min
Delta R.T.: 0.008 min
Response: 1514051
Conc: 5.11 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 7.247 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 6.028 min
Delta R.T.: 0.006 min
Response: 303896
Conc: 1.07 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.829 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#35 AR-1260-5

R.T.: 6.865 min
Delta R.T.: 0.069 min
Response: 1061872
Conc: 1.86 ng/ml




#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 7.247 min
Response: 0
Conc: N.D.



#36 AR-1262-1


R.T.: 6.247 min
Delta R.T.: -0.007 min
Response: 328242
Conc: 0.92 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T.: 7.828 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#37 AR-1262-2

R.T.: 6.865 min
Delta R.T.: 0.068 min
Response: 1061872
Conc: 1.67 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.134 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.092 min
Delta R.T.: -0.010 min
Response: 289484
Conc: 1.16 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.130 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114




#41 AR-1268-1

R.T.: 7.092 min
Delta R.T.: -0.011 min
Response: 289484
Conc: 0.39 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 8.438 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 7.393 min
Delta R.T.: -0.001 min
Response: 324379
Conc: 0.56 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.226 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :
AIBLK114



#45 AR-1268-5

R.T.: 8.004 min
Delta R.T.: -0.002 min
Response: 1460109
Conc: 0.80 ng/ml