

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
Data File : PR069371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 14:27
Operator : AJ\MA
Sample : P5066-02
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
E28M1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/20/2024
Supervised By :Ankita Jodhani 12/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 22:29:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) AR-1242-1 (L4)
R.T. Response Conc
4.86 2553584 96.68
4.88 3619883 91.53
4.95 2199711 89.15
5.05 2218382 115.31
5.84 4444063 216.74

(16) AR-1242-1 #2 (L4)
R.T. Response Conc
4.09 10320110 61.73
4.11 19062971 80.23
4.29 12402919 97.29
4.38 15223987 125.65
4.93 34909825 242.06

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
 Data File : PR069371.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 14:27
 Operator : AJ\MA
 Sample : P5066-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28M1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/20/2024
 Supervised By :Ankita Jodhani 12/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 08:52:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) AR-1242-1 (L4)
 R.T. Response Conc
 4.86 2553584 96.68
 4.88 3619883 91.53
 4.95 2199711 89.15
 5.05 2218382 115.31
 5.84 4444063 216.74

(16) AR-1242-1 #2 (L4)
 R.T. Response Conc
 4.09 11336107 67.80
 4.11 18261960 76.86
 4.29 12402919 97.29
 4.38 15223987 125.65
 4.93 34909825 242.06

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
Data File : PR069371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Dec 2024 14:27
Operator : AJ\MA
Sample : P5066-02
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
E28M1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/20/2024
Supervised By :Ankita Jodhani 12/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 22:29:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.58 5886183 142.62
6.86 9946849 208.46
7.25 4138166 112.95
7.49 9188037 220.89
7.83 10192482 133.59

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.66 38153556 150.60
5.87 42983151 145.03
6.02 37259574 131.47
6.53 21939898 90.84
6.80 51932261 91.05

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR120624\
 Data File : PR069371.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Dec 2024 14:27
 Operator : AJ\MA
 Sample : P5066-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28M1

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/20/2024
 Supervised By :Ankita Jodhani 12/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 08:52:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.58 5886183 142.62
 6.86 9946849 208.46
 7.25 4138166 112.95
 7.49 4651470 111.83
 7.83 10192482 133.59

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.66 38153556 150.60
 5.87 42983151 145.03
 6.02 37259574 131.47
 6.53 21939898 90.84
 6.80 51932261 91.05

(+) = Expected Retention Time