

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121119\
 Data File : PR043702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2019 10:56
 Operator : SM\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 ECD_R
 LabSampleId :
 AR1660363

Manual Integrations
 APPROVED

Ankita
 12/12/2019 11:09:07 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 16:25:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:51:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)
 R.T. Response Conc
 5.89 81589082 421.38
 5.91 116286175 424.69
 5.98 72942538 442.92
 6.08 61621740 447.45
 6.37 60310519 436.47

(3) AR-1016-1 #2 (L1)
 R.T. Response Conc
 5.06 147757643 362.42
 5.08 207356829 371.23
 5.25 111419235 372.94
 5.29 90496346 376.23
 5.51 123254774 388.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121119\
Data File : PR043702.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2019 10:56
Operator : SM\AJ
Sample : AR1660CCC400
Misc :
ALS Vial : 3 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1660363

Manual Integrations
APPROVED

Ankita
12/12/2019 11:09:07 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 11 16:25:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 07 03:51:56 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(3) AR-1016-1 (L1)
R.T. Response Conc
5.89 78030568 403.00
5.91 112307724 410.16
5.98 66703546 405.04
6.08 55897573 405.88
6.37 55675549 402.92

(3) AR-1016-1 #2 (L1)
R.T. Response Conc
5.06 147757643 362.42
5.08 207356829 371.23
5.25 111419235 372.94
5.29 90496346 376.23
5.51 123254774 388.17

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121119\
 Data File : PR043702.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Dec 2019 10:56
 Operator : SM\AJ
 Sample : AR1660CCC400
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1660363

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 11 16:25:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120419CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 07 03:51:56 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Manual Integrations
 APPROVED

Ankita
 12/12/2019 11:09:07 AM

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.713	3.961	107.2E6	135.0E6	19.291	17.480
2) SA Decachlor...	10.625	9.058	164.8E6	356.3E6	43.753	36.004
Target Compounds						
3) L1 AR-1016-1	5.890	5.057	78030568	147.8E6	403.001m	362.422
4) L1 AR-1016-2	5.913	5.076	112.3E6	207.4E6	410.163m	371.229
5) L1 AR-1016-3	5.976	5.255	66703546	111.4E6	405.036m	372.941
6) L1 AR-1016-4	6.077	5.294	55897573	90496346	405.881m	376.226
7) L1 AR-1016-5	6.370	5.511	55675549	123.3E6	402.923m	388.170
31) L7 AR-1260-1	7.498	6.548	112.1E6	264.1E6	449.944	389.842
32) L7 AR-1260-2	7.754	6.736	133.4E6	383.0E6	449.038	395.307
33) L7 AR-1260-3	8.115	6.891	100.4E6	313.5E6	453.351	401.490
34) L7 AR-1260-4	8.353	7.364	116.9E6	250.5E6	464.883	394.813
35) L7 AR-1260-5	8.691	7.605	231.3E6	670.6E6	479.654	393.951

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Handwritten signature: AJ
 12/13/19