

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
Data File : PR069452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2024 15:11
Operator : AJ\MA
Sample : P5234-08
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
C0AR2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:50:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.631min 16.491 ng/ml

response 17250508

(1) Tetrachloro-m-xylene #2 (SA)

2.970min 18.596 ng/ml

response 115774640

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
Data File : PR069452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2024 15:11
Operator : AJ\MA
Sample : P5234-08
Misc :
ALS Vial : 16 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
C0AR2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:50:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.631min 28.390 ng/ml m

response 29697753

(1) Tetrachloro-m-xylene #2 (SA)

2.970min 18.596 ng/ml

response 115774640



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
Data File : PR069452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2024 15:11
Operator : AJ\MA
Sample : P5234-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
C0AR2

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:50:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
R.T. Response Conc
6.57 24362705 590.31
6.85 68130167 1427.86
7.25 18034837 492.27
7.49 23234579 558.59
7.83 35427139 464.34

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.65 226695750 894.82
5.86 195187964 658.59
6.02 134203290 473.54
6.53 61878248 256.20
6.79 192367200 337.27

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
Data File : PR069452.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Dec 2024 15:11
Operator : AJ\MA
Sample : P5234-08
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AR2

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 01:50:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
R.T. Response Conc
6.57 24362705 590.31
6.85 68130167 1427.86
7.25 18034837 492.27
7.49 12229723 294.02
7.83 35427139 464.34
(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.65 170364159 672.47
5.86 195187964 658.59
6.02 134203290 473.54
6.53 61878248 256.20
6.79 192367200 337.27