

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
Data File : PR069490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2024 01:14
Operator : AJ\MA
Sample : P5232-16
Misc :
ALS Vial : 48 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
E28T0

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 12 02:01:14 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Reviewed By :Yogesh Patel 12/12/2024
Supervised By :Ankita Jodhani 12/16/2024

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.57	1480274	35.87
6.86	1570738	32.92
7.24	1974760	53.90
7.49	3006681	72.28
7.83	3296290	43.20

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.65	8007845	31.61
5.86	14165800	47.80
6.02	5134261	18.12
6.53	8881538	36.77
6.79	17613052	30.88

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121124\
 Data File : PR069490.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2024 01:14
 Operator : AJ\MA
 Sample : P5232-16
 Misc :
 ALS Vial : 48 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28T0

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/12/2024
 Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 12 02:01:14 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.57	1480274	35.87
6.86	1570738	32.92
7.24	1974760	53.90
7.49	1895506	45.57
7.83	3296290	43.20

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.65	8007845	31.61
5.86	11918259	40.21
6.02	8843584	31.20
6.53	8881538	36.77
6.79	17613052	30.88