

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121220\
 Data File : PR048928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Dec 2020 09:02
 Operator : DD\AJ
 Sample : AR1242CCC400
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242343

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 14 00:31:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.695	3.847	48554204	90242250	18.373	21.215
2) SA Decachlor...	10.599	8.921	81145648	297.2E6	30.923	34.672
Target Compounds						
16) L4 AR-1242-1	5.870	4.932	29919612	25356176	363.990	408.277
17) L4 AR-1242-2	5.894	4.951	41169088	53486370	358.135	403.067
18) L4 AR-1242-3	5.956	5.128	25436495	29126506	357.703	398.696
19) L4 AR-1242-4	6.056	5.211	20871126	17389518	357.970	409.793
20) L4 AR-1242-5	6.794	5.735	22433189	30747385	339.768	473.214 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121220\
Data File : PR048928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Dec 2020 09:02
Operator : DD\AJ
Sample : AR1242CCC400
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242343

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 14 00:31:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 01 06:11:17 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

