

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121224\
Data File : PR069560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2024 00:07
Operator : AJ\MA
Sample : P5234-10DL 10X
Misc :
ALS Vial : 49 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AR4DL

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 01:51:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(16) AR-1242-1 (L4)
R.T. Response Conc
4.85 6573744 248.89
4.88 11044752 279.26
4.95 2000761 81.09
5.04 3843735 199.80
5.83 20060510 978.37

(16) AR-1242-1 #2 (L4)
R.T. Response Conc
4.09 12496930 74.74
4.11 46578673 196.03
4.29 35995764 282.36
4.38 94388335 779.02
4.93 302827184 2099.79

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121224\
Data File : PR069560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2024 00:07
Operator : AJ\MA
Sample : P5234-10DL 10X
Misc :
ALS Vial : 49 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AR4DL

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 01:51:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(16) AR-1242-1 (L4)

R.T.	Response	Conc
4.85	6573744	248.89
4.88	11044752	279.26
4.95	2000761	81.09
5.04	3843735	199.80
5.83	20060510	978.37

(16) AR-1242-1 #2 (L4)

R.T.	Response	Conc
4.09	20659121	123.56
4.11	59688073	251.21
4.29	35995764	282.36
4.38	94388335	779.02
4.93	302827184	2099.79

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121224\
 Data File : PR069560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2024 00:07
 Operator : AJ\MA
 Sample : P5234-10DL 10X
 Misc :
 ALS Vial : 49 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 C0AR4DL

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 01:51:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 12/13/2024
 Supervised By :Ankita Jodhani 12/16/2024

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.57 31940743 773.92
 6.85 38333693 803.39
 7.24 14416972 393.52
 7.47 28122538 676.11
 7.82 21371447 280.11
 (31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.65 239829672 946.67
 5.86 231620507 781.52
 6.01 209812513 740.33
 6.52 60030489 248.55
 6.79 142312296 249.51

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121224\
Data File : PR069560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2024 00:07
Operator : AJ\MA
Sample : P5234-10DL 10X
Misc :
ALS Vial : 49 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AR4DL

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/13/2024
Supervised By :Ankita Jodhani 12/16/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 01:51:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)	R.T.	Response	Conc
6.57	31940743	773.92	
6.85	38333693	803.39	
7.24	14416972	393.52	
7.48	12179777	292.82	
7.82	21371447	280.11	

(31) AR-1260-1 #2 (L7)	R.T.	Response	Conc
5.65	239829672	946.67	
5.86	231620507	781.52	
6.01	209812513	740.33	
6.52	60030489	248.55	
6.79	142312296	249.51	