

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121318\
 Data File : PR034584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Dec 2018 14:35
 Operator : SM\SJ
 Sample : J6214-33RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 13 15:12:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 07 02:02:36 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.440	3.518	39313301	75800539	21.779	20.587
2) SA Decachlor...	10.127	8.420	19313897	40097953	9.840	9.801
Target Compounds						
7) L1 AR-1016-5	0.000	5.055	0	9068655	N.D.	82.306 #
15) L3 AR-1232-5	0.000	5.055	0	9068655	N.D.	230.919 #
20) L4 AR-1242-5	6.455	0.000	23917	0	0.521	N.D. #
24) L5 AR-1248-4	6.455	5.055	23917	9068655	0.299	61.171 #
25) L5 AR-1248-5	6.455	0.000	23917	0	0.317	N.D. #
26) L6 AR-1254-1	6.455	0.000	23917	0	0.294	N.D. #
28) L6 AR-1254-3	7.046	5.892	2516580	20233495	18.248	56.001 #
31) L7 AR-1260-1	0.000	6.036	0	1926875	N.D.	7.983 #
32) L7 AR-1260-2	0.000	6.210	0	6513151	N.D.	21.490 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121318\
Data File : PR034584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Dec 2018 14:35
Operator : SM\SJ
Sample : J6214-33RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 13 15:12:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR120718.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 07 02:02:36 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.440 min
Delta R.T.: 0.008 min
Response: 39313301
Conc: 21.78 ng/ml

Instrument :
ECD_R
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.518 min
Delta R.T.: -0.001 min
Response: 75800539
Conc: 20.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.127 min
Delta R.T.: 0.004 min
Response: 19313897
Conc: 9.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.420 min
Delta R.T.: -0.005 min
Response: 40097953
Conc: 9.80 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T. : 6.072 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :




#7 AR-1016-5

R.T.: 5.055 min
Delta R.T.: 0.031 min
Response: 9068655
Conc: 82.31 ng/ml




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.869 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.055 min
Delta R.T.: 0.030 min
Response: 9068655
Conc: 230.92 ng/ml



#20 AR-1242-5


R.T.: 6.455 min
Delta R.T.: -0.057 min
Response: 23917
Conc: 0.52 ng/ml

Instrument :
ECD_R
ClientSampleId :




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.369 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.455 min
Delta R.T.: -0.018 min
Response: 23917
Conc: 0.30 ng/ml



#24 AR-1248-4


R.T.: 5.055 min
Delta R.T.: 0.030 min
Response: 9068655
Conc: 61.17 ng/ml



#25 AR-1248-5


R.T.: 6.455 min
Delta R.T.: -0.058 min
Response: 23917
Conc: 0.32 ng/ml

Instrument :
ECD_R
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.409 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.455 min
Delta R.T.: 0.007 min
Response: 23917
Conc: 0.29 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.369 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.046 min
Delta R.T.: 0.016 min
Response: 2516580
Conc: 18.25 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 5.892 min
Delta R.T.: -0.020 min
Response: 20233495
Conc: 56.00 ng/ml




#31 AR-1260-1

R.T.: 0.000 min
Exp R.T. : 7.192 min
Response: 0
Conc: N.D.



#31 AR-1260-1


R.T.: 6.036 min
Delta R.T.: -0.003 min
Response: 1926875
Conc: 7.98 ng/ml



#32 AR-1260-2

R.T.: 0.000 min
Exp R.T.: 7.446 min
Response: 0
Conc: N.D.

Instrument :
ECD_R
ClientSampleId :



#32 AR-1260-2

R.T.: 6.210 min
Delta R.T.: -0.014 min
Response: 6513151
Conc: 21.49 ng/ml