


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR121324\  
Data File : PR069601.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 13 Dec 2024 12:39  
Operator : AJ\MA  
Sample : P5234-01DL2 60X  
Misc :  
ALS Vial : 77 Sample Multiplier: 1

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Dec 14 01:37:40 2024  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR112124CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri Nov 22 04:30:58 2024  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

3.629min 0.713 ng/ml

response 746324

(1) Tetrachloro-m-xylene #2 (SA)

2.966min 0.346 ng/ml m

response 2151534