

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121420\
 Data File : PR048957.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2020 13:11
 Operator : DD\AJ
 Sample : L4999-09
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0BX5

Manual Integrations
 APPROVED

Ankita
 12/15/2020 3:14:10 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 01:22:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.688	3.843	33253638	59620285	12.583	14.016
2) SA Decachlor...	10.595	8.926	57127983	211.0E6	21.771	24.617
Target Compounds						
26) L6 AR-1254-1	6.723	5.736	1900075	1678488	17.341	14.025m
27) L6 AR-1254-2	6.941	5.882	3098593	2506925	18.600	20.308
28) L6 AR-1254-3	7.312	6.289	3118418	5631546	17.857	22.320m
29) L6 AR-1254-4	7.598	6.519	2177680	4869421	16.122	15.846
30) L6 AR-1254-5	8.017	6.939	2961665	10098803	20.665	23.126m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121420\
Data File : PR048957.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2020 13:11
Operator : DD\AJ
Sample : L4999-09
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
C0BX5

Manual Integrations
APPROVED

Ankita
12/15/2020 3:14:10 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 01:22:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 01 06:11:17 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

