

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121423\
 Data File : PR064684.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Dec 2023 19:45
 Operator : AJ\MA
 Sample : AR1248ICC400
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 15 01:03:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Dec 15 01:02:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.702	3.007	116.9E6	110.8E6	20.000	20.000
2) SA Decachlor...	9.720	8.384	87844346	104.1E6	40.000	40.000
Target Compounds						
21) L5 AR-1248-1	4.941	4.155	43994924	44313569	400.000	400.000
22) L5 AR-1248-2	5.237	4.408	62214176	61157524	400.000	400.000
23) L5 AR-1248-3	5.451	4.451	72410681	62716218	400.000	400.000
24) L5 AR-1248-4	5.890	4.632	77315748	75685986	400.000	400.000
25) L5 AR-1248-5	5.929	5.054	77897024	75229187	400.000	400.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121423\
Data File : PR064684.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Dec 2023 19:45
Operator : AJ\MA
Sample : AR1248ICC400
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 15 01:03:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 15 01:02:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#21 AR-1248-1


R.T.: 4.941 min
Delta R.T.: 0.000 min
Response: 43994924
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#21 AR-1248-1

R.T.: 4.155 min
Delta R.T.: 0.000 min
Response: 44313569
Conc: 400.00 ng/ml




#22 AR-1248-2

R.T.: 5.237 min
Delta R.T.: 0.000 min
Response: 62214176
Conc: 400.00 ng/ml



#22 AR-1248-2


R.T.: 4.408 min
Delta R.T.: 0.000 min
Response: 61157524
Conc: 400.00 ng/ml



#23 AR-1248-3


R.T.: 5.451 min
Delta R.T.: 0.000 min
Response: 72410681
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :




#23 AR-1248-3

R.T.: 4.451 min
Delta R.T.: 0.000 min
Response: 62716218
Conc: 400.00 ng/ml




#24 AR-1248-4

R.T.: 5.890 min
Delta R.T.: 0.000 min
Response: 77315748
Conc: 400.00 ng/ml



#24 AR-1248-4


R.T.: 4.632 min
Delta R.T.: 0.000 min
Response: 75685986
Conc: 400.00 ng/ml



#25 AR-1248-5

R.T.: 5.929 min
Delta R.T.: 0.000 min
Response: 77897024
Conc: 400.00 ng/ml

Instrument :
ECD_R
ClientSampleId :



#25 AR-1248-5

R.T.: 5.054 min
Delta R.T.: 0.000 min
Response: 75229187
Conc: 400.00 ng/ml