

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121523\
Data File : PR064726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 14:34
Operator : AJ\MA
Sample : 05794-22
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
BGRG3

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2023
Supervised By :Ankita Jodhani 12/19/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 04:59:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 15 04:19:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(31) AR-1260-1 (L7)
R.T. Response Conc
6.68 1508907226 6038.28
6.95 18580859168 65002.41
7.36 1050978707 4956.72
7.60 1217132735 5295.29
7.94 2823818476 6581.03

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.75 1622114147 5613.31
5.96 2518894674 7250.05
6.12 1840405724 5591.35
6.63 1113192900 4127.17
6.90 2993538509 4790.31

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121523\
Data File : PR064726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Dec 2023 14:34
Operator : AJ\MA
Sample : 05794-22
Misc :
ALS Vial : 27 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
BGRG3

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2023
Supervised By :Ankita Jodhani 12/19/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 16 04:59:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121423CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Dec 15 04:19:57 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.68 1508907226 6038.28
6.95 18580859168 65002.41
7.36 1050978707 4956.72
7.60 1358469581 5910.19
7.94 2823818476 6581.03

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.75 1622114147 5613.31
5.96 2518894674 7250.05
6.12 1840405724 5591.35
6.63 1113192900 4127.17
6.90 2993538509 4790.31