

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121624\
 Data File : PR069692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Dec 2024 23:19
 Operator : AJ\MA
 Sample : P5264-08MSD
 Misc :
 ALS Vial : 42 Sample Multiplier: 1


Instrument :
 ECD_R
 ClientSampleId :
 C0AS6MSD

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 01:25:08 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 12/18/2024
 Supervised By :Ankita Jodhani 12/18/2024

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.57	20493229	496.55
6.85	24629853	516.19
7.24	14333937	391.25
7.48	20442373	491.46
7.82	30805653	403.76

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.65	121845503	480.95
5.86	140242695	473.20
6.01	128084235	451.95
6.52	85143605	352.53
6.79	194713990	341.38

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121624\
Data File : PR069692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Dec 2024 23:19
Operator : AJ\MA
Sample : P5264-08MSD
Misc :
ALS Vial : 42 Sample Multiplier: 1


Instrument :
ECD_R
ClientSampleId :
C0AS6MSD

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2024
Supervised By :Ankita Jodhani 12/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 01:25:08 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)

R.T.	Response	Conc
6.57	20493229	496.55
6.85	24629853	516.19
7.24	14333937	391.25
7.48	18128475	435.83
7.82	30805653	403.76

(31) AR-1260-1 #2 (L7)

R.T.	Response	Conc
5.65	121845503	480.95
5.86	140242695	473.20
6.01	128084235	451.95
6.52	85143605	352.53
6.79	194713990	341.38