

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121724\
 Data File : PR069726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2024 18:08
 Operator : AJ\MA
 Sample : P5265-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28Y8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2024
 Supervised By :Ankita Jodhani 12/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 02:22:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
 R.T. Response Conc
 5.76 2046689 56.95
 6.00 4652230 84.54
 6.39 5033291 88.10
 6.70 3613711 88.89
 7.15 7710488 180.19

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 4.92 14337207 48.05
 5.08 22220437 86.76
 5.51 19332945 48.93
 5.75 11951886 48.80
 6.20 35836185 102.69

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121724\
 Data File : PR069726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2024 18:08
 Operator : AJ\MA
 Sample : P5265-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28Y8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2024
 Supervised By :Ankita Jodhani 12/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 02:22:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
 R.T. Response Conc
 5.76 723853 20.14
 6.00 1960692 35.63
 6.39 2493569 43.65
 6.70 1709553 42.05
 7.15 6138803 143.46

(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 4.92 11082509 37.14
 5.08 11658777 45.52
 5.51 14179885 35.89
 5.75 7224705 29.50
 6.20 35836185 102.69

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121724\
Data File : PR069726.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2024 18:08
Operator : AJ\MA
Sample : P5265-03
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
E28Y8

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 12/18/2024
Supervised By :Ankita Jodhani 12/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 02:22:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 04:30:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
R.T. Response Conc
6.57 5385873 130.50
6.85 6544621 137.16
7.24 6510095 177.70
7.48 7907677 190.11
7.82 11200761 146.81

(31) AR-1260-1 #2 (L7)
R.T. Response Conc
5.65 40101719 158.29
5.85 47340245 159.73
6.01 24510212 86.48
6.52 31035205 128.50
6.79 65868474 115.48

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121724\
 Data File : PR069726.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2024 18:08
 Operator : AJ\MA
 Sample : P5265-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1



Instrument :
 ECD_R
 ClientSampleId :
 E28Y8

Manual IntegrationsAPPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 02:22:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR12124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 04:30:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Reviewed By :Yogesh Patel 12/18/2024
 Supervised By :Ankita Jodhani 12/18/2024

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(31) AR-1260-1 (L7)
 R.T. Response Conc
 6.57 4092234 99.15
 6.85 5518993 115.67
 7.24 5659999 154.49
 7.48 6631496 159.43
 7.82 10231480 134.10

(31) AR-1260-1 #2 (L7)
 R.T. Response Conc
 5.65 22860784 90.24
 5.85 31984356 107.92
 6.01 24510212 86.48
 6.52 31035205 128.50
 6.79 65868474 115.48