

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR121818CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 18 01:15:04 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034721.D 1600 =PR034723.D 800 =PR034722.D
 200 =PR034720.D 100 =PR034719.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.894	1.870	1.829	1.931	2.119	1.929	E6	5.84
2) SA	Decachlorobiphenyl	1.999	1.846	1.877	2.156	2.426	2.061	E6	11.54
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1	4.789	4.332	4.429	5.067	5.645	4.852	E4	10.94
22) L5	AR-1248-2	6.467	5.828	6.008	6.878	7.855	6.607	E4	12.24
23) L5	AR-1248-3	7.342	6.711	6.872	7.770	8.663	7.471	E4	10.50
24) L5	AR-1248-4	0.867	0.787	0.811	0.927	1.076	0.893	E5	12.91
25) L5	AR-1248-5	8.164	7.437	7.662	8.573	9.998	8.367	E4	12.11
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR034721.D 1600 =PR034723.D 800 =PR034722.D
 200 =PR034720.D 100 =PR034719.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.742	3.738	3.612	3.776	4.112	3.796	E6	4.93
2) SA	Decachlorobiphenyl	4.565	4.435	4.417	4.793	5.350	4.712	E6	8.21
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR121818CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 18 01:15:04 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034721.D 1600 =PR034723.D 800 =PR034722.D
 200 =PR034720.D 100 =PR034719.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1	0.956	0.887	0.898	0.995	1.140	0.975 E5	10.44
22) L5	AR-1248-2	1.254	1.159	1.176	1.313	1.504	1.281 E5	10.85
23) L5	AR-1248-3	1.291	1.196	1.213	1.351	1.548	1.320 E5	10.77
24) L5	AR-1248-4	1.605	1.511	1.522	1.666	1.923	1.645 E5	10.17
25) L5	AR-1248-5	1.638	1.557	1.564	1.690	1.920	1.674 E5	8.86
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range