

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR121818CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 18 01:02:32 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034726.D 1600 =PR034728.D 800 =PR034727.D
 200 =PR034725.D 100 =PR034724.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.983	1.747	1.868	1.930	1.984	1.902	E6	5.20
2) SA	Decachlorobiphenyl	1.958	1.828	1.846	2.168	2.377	2.036	E6	11.51
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1	8.088	7.041	7.335	8.793	9.595	8.171	E4	12.83
27) L6	AR-1254-2	1.248	1.114	1.150	1.367	1.510	1.278	E5	12.74
28) L6	AR-1254-3	1.305	1.213	1.225	1.435	1.571	1.350	E5	11.28
29) L6	AR-1254-4	1.025	0.940	0.957	1.134	1.247	1.061	E5	12.18
30) L6	AR-1254-5	1.049	0.991	0.994	1.115	1.208	1.071	E5	8.52
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR034726.D 1600 =PR034728.D 800 =PR034727.D
 200 =PR034725.D 100 =PR034724.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.917	3.509	3.731	3.819	4.025	3.800	E6	5.17
2) SA	Decachlorobiphenyl	4.504	4.430	4.367	4.888	5.320	4.702	E6	8.52
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR121818CLP.M
 Title : GC EXTRACTABLES
 Last Update : Tue Dec 18 01:02:32 2018
 Response Via : Initial Calibration

Calibration Files

400 =PR034726.D 1600 =PR034728.D 800 =PR034727.D
 200 =PR034725.D 100 =PR034724.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.418	2.181	2.236	2.584	2.815	2.447 E5	10.62
27) L6	AR-1254-2	2.072	1.906	1.941	2.249	2.467	2.127 E5	10.96
28) L6	AR-1254-3	3.455	3.323	3.317	3.741	4.029	3.573 E5	8.60
29) L6	AR-1254-4	2.251	2.182	2.175	2.482	2.718	2.362 E5	9.94
30) L6	AR-1254-5	3.038	2.987	2.948	3.359	3.615	3.189 E5	9.03
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range