

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121818\
 Data File : PR034735.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Dec 2018 21:56
 Operator : SM\SJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1254316

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 18 02:04:40 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 18 01:56:32 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.426	3.511	34128741	66636504	17.547	19.115
2)	SA Decachlor...	10.109	8.412	82126884	189.9E6	41.775	43.184
Target Compounds							
26)	L6 AR-1254-1	6.438	5.358	31176237	93470507	381.571	381.990
27)	L6 AR-1254-2	6.655	5.502	49231876	81530476	385.296	383.331
28)	L6 AR-1254-3	7.021	5.899	53082918	139.7E6	393.288	390.914
29)	L6 AR-1254-4	7.305	6.125	42300734	92967311	398.771	393.612
30)	L6 AR-1254-5	7.723	6.536	44307045	128.2E6	413.530	401.941

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121818\
Data File : PR034735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Dec 2018 21:56
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
AR1254316

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 18 02:04:40 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121818CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 18 01:56:32 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

