

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121820\
Data File : PR049002.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Dec 2020 20:04
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_R
LabSampleId :
AR1254302

Manual Integrations
APPROVED

Ankita
12/21/2020 3:13:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 19 01:52:40 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Dec 01 06:11:17 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



QEdit

(26) AR-1254-1 (L6)

R.T.	Response	Conc
6.72	38942679	355.42
6.94	58865453	353.35
7.31	60604712	347.05
7.60	48056513	355.78
8.02	51609258	360.11

(26) AR-1254-1 #2 (L6)

R.T.	Response	Conc
5.74	47286937	395.12
5.88	46515044	376.81
6.29	99544597	394.54
6.52	157631252	512.95
6.94	189233897	433.34

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121820\
 Data File : PR049002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2020 20:04
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 ECD_R
 LabSampleId :
 AR1254302

Manual Integrations
 APPROVED

Ankita
 12/21/2020 3:13:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 01:52:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 6.72 38942679 355.42
 6.94 58865453 353.35
 7.31 60604712 347.05
 7.60 48056513 355.78
 8.02 51609258 360.11
 (26) AR-1254-1 #2 (L6)
 R.T. Response Conc
 5.74 47286937 395.12
 5.88 46515044 376.81
 6.29 99544597 394.54
 6.52 134592686 437.98
 6.94 189233897 433.34

QEdit

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121820\
 Data File : PR049002.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Dec 2020 20:04
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 AR1254302

Manual Integrations
 APPROVED

Ankita
 12/21/2020 3:13:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 19 01:52:40 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR113020CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Dec 01 06:11:17 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.690	3.845	45151540	85253453	17.086	20.042
2) SA Decachlor...	10.593	8.931	88770594	342.5E6	33.829	39.952
Target Compounds						
26) L6 AR-1254-1	6.724	5.736	38942679	47286937	355.418	395.123
27) L6 AR-1254-2	6.942	5.884	58865453	46515044	353.347	376.815
28) L6 AR-1254-3	7.312	6.291	60604712	99544597	347.049	394.536
29) L6 AR-1254-4	7.598	6.520	48056513	134.6E6	355.780	437.979m
30) L6 AR-1254-5	8.019	6.941	51609258	189.2E6	360.107	433.340

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AS
 12/24/20