

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121922\
 Data File : PR058586.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Dec 2022 16:02
 Operator : YP\AJ
 Sample : N6128-04DL 20X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 A43T3DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 20 13:32:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Dec 15 03:47:13 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.677	2.914	126.0E6	4263914	41.760	1.283 #
2)	SA Decachlor...	9.659	8.157	6070547	8068073	3.314	2.957
Target Compounds							
16)	L4 AR-1242-1	4.925	4.027	651.4E6	1065.8E6	8654.296	10940.883 #
17)	L4 AR-1242-2	4.948	4.045	1140.1E6	1916.2E6	11144.288	14326.899 #
18)	L4 AR-1242-3	5.012	4.223	401.2E6	852.1E6	6005.641	11862.386 #
19)	L4 AR-1242-4	5.115	4.315	493.1E6	647.9E6	9148.095	9515.310
20)	L4 AR-1242-5	5.909	4.859	171.1E6	410.7E6	3248.793	4598.869 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR121922\
Data File : PR058586.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Dec 2022 16:02
Operator : YP\AJ
Sample : N6128-04DL 20X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
A43T3DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 20 13:32:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR121422CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Dec 15 03:47:13 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

